



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Justin Weikert</i>		
Location <i>Cooperstown</i>		Intoxilyzer® Serial Number <i>80-005362</i>
Gas Lot Number <i>16512040A2</i>	Gas Cylinder Number <i>30</i>	Gas Expiration Date <i>09/01/2014</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. *0.083* AC
  - B. *0.082* AC
  - C. *0.083* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>2/22/13</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>25 Feb 2013</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005362  
Location = COOP                    8164.13.00 06/09  
02/22/2013                            14:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:06
02 Std. Gas	0.083	14:06
03 Room Air	0.000	14:06
04 Std. Gas	0.082	14:07
05 Room Air	0.000	14:07
06 Std. Gas	0.083	14:08
07 Room Air	0.000	14:08

Lot No = 16512080A2  
Cyl No = 30  
Exp Date = 08/01/2014  
County = 20                            Oper No. = 011017

  
Operator Signature  
JUSTIN WALLERICH

Remarks: