



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>William Petty</u>		
Location: <u>GFAE</u>		
Intoxilyzer® Serial Number: <u>80-002667</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>39</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>081</u> %		
Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

William Petty  
Chemical Test Operator Signature

23 Feb 11  
Date

Deb Keshun  
Reviewed By

22 Mar 2011  
Date

80-002667  
p-88 of 89  
SFN


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = GFAF      8164.13.00 06/09  
02/23/2011      08:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:43
02 Std. Gas	0.080	08:44
03 Room Air	0.000	08:44
04 Std. Gas	0.081	08:44
05 Room Air	0.000	08:45
06 Std. Gas	0.081	08:45
07 Room Air	0.000	08:46

Lot No = 659358  
Cyl No = 39  
Exp Date = 09/03/2011  
County = 18      Oper No. = 120011



Operator Signature  
WILLIAM R. PETTY

Remarks:

Form 106-I8000

80-002667  
P. 89 of 89  
RM