



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David Feland</u>		
Location: <u>Kidder So.</u>		
Intoxilyzer® Serial Number: <u>80-004186</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>137^{DK}37</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.083</u> %		
Test 2: <u>0.082</u> %		
Test 3: <u>0.082</u> %		
(ATTACH TEST RECORD)		

David Feland
Chemical Test Operator Signature

24 Feb 11
Date

Deb Kashur
Reviewed By

01 Mar 2011
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
02/24/2011 14:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:03
02 Std. Gas	0.083	14:04
03 Room Air	0.000	14:04
04 Std. Gas	0.082	14:04
05 Room Air	0.000	14:05
06 Std. Gas	0.082	14:05
07 Room Air	0.000	14:06

Lot No = 659357
Cyl No = 37
Exp Date = 09/03/2011
County = 22 Oper No. = 119404



Operator Signature
DAVID R. FELAND

Remarks:

Form 106-I8000

80-004186
p. 100 of 100
Jm