



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Oase, Denver L</i>		
Location <i>Mott</i>	Intoxilyzer® Serial Number <i>80-004200</i>	
Gas Lot Number <i>19615080 AS</i>	Gas Cylinder Number <i>25</i>	Gas Expiration Date <i>09/05/2017</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.082* AC
 - B. *0.081* AC
 - C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>2/24/2016</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Shanaver</i>	Date <i>01 Mar 2016</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = MOTT 8164.13.00 06/09
02/24/2016 13:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:01
02 Std. Gas	0.082	14:02
03 Room Air	0.000	14:02
04 Std. Gas	0.081	14:03
05 Room Air	0.000	14:03
06 Std. Gas	0.081	14:03
07 Room Air	0.000	14:04


Lot No = 19615080A5

Cyl No = 25

Exp Date = 09/05/2017

County = 21

Oper No. = 130994

 355

Operator Signature
DENVER OASE

Remarks:

Form 106-I8000