

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004199  
Location = ROLL      8164.13.00 06/09  
02/25/2010      20:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:03
02 Std. Gas	0.083	20:04
03 Room Air	0.000	20:04
04 Std. Gas	0.082	20:05
05 Room Air	0.000	20:05
06 Std. Gas	0.082	20:05
07 Room Air	0.000	20:06

Lot No = 659358  
Cyl No = 11  
Exp Date = 09/03/2011  
County = 40      Oper No. = 049904



Operator Signature  
STEVEN R. FISCHER

Remarks:

Form 106-I8000

80-004199  
p. 75 of 75  
JFH