

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004191

Location: TOXL

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5144</u> Lot No. <u>    </u> Rep AC <u>0.021</u>   | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR3847</u> Lot No. <u>    </u> Rep AC <u>0.294</u>                                | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3452</u> Lot No. <u>    </u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>15410080A1</u>	Cyl No. <u>22</u>	Exp. Date <u>07/01/2012</u>
Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.081</u>		

L. Tank pressure, Level 3 Func D Sub G P: 806 psi Reg 800 psi

M. Remarks/Maintenance record: Annual inspection w/software 8/04.13

25 Feb 2011  
Date

Deb Kashur  
Field Inspector's Signature

28 Feb 2011  
Date

Margy Plam  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000       SN 80-004191  
Location = TOXL               8164.13.00 06/09  
02/25/2011                       14:59

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%^&\*()\_+?

Current Instrument Setup


Data Entry Mode:           Enabled  
Start Test Sequence:       DABACABA  
Display Prelim Rslt?       Yes  
Display Third Digit?       Yes  
Inhib Printer(Y/N)?       No  
Display Volume?            No  
Disable On Memfull?       Yes  
# of Print Copies?         1  
Select Std (D/W/I)?       Dry  
Standard Value?            0.080  
Standard Lot #?            15410080A1  
Standard Cyl #?            22  
Standard Expiration?       07/01/2012  
Oper No?                   777777

Flow Cal. Date:            01/28/2010  
Slope                       738  
Intercept                  -777574

IR Calibration Date:       01/28/2010  
                              3um                       9um


-----  
0th Coef(\*100):       -23128               -33609  
1st Coef(\*100):       272485               131724  
2nd Coef(\*100):       1490                   1056  
H2O adj(mg/l\*10k):   432                    614

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DEB KASHUR

Remarks:



Form 106-I8000

80-004191  
P. 69 of 76  
VM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = TOXL      8164.13.00 06/09  
02/25/2011      15:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:39
02 Std. Sol.	0.021	15:39
03 Room Air	0.000	15:40
04 Std. Sol.	0.021	15:41
05 Room Air	0.000	15:41
06 Std. Sol.	0.021	15:42
07 Room Air	0.000	15:42

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC      0.021AC*

Form 106-I8000

*80-004191  
p. 70 of 76  
FKM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = TOXL      8164.13.00 06/09  
02/25/2011      15:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:50
02 Std. Sol.	0.296	15:51
03 Room Air	0.000	15:52
04 Std. Sol.	0.296	15:53
05 Room Air	0.000	15:53
06 Std. Sol.	0.296	15:54
07 Room Air	0.000	15:54

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.294 AC*

Form 106-I8000

80-004191  
p. 71 of 76  
VFM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004191  
Location = TOXL                8164.13.00 06/09  
02/25/2011                        16:11

Test	AC	Time
01 Room Air	0.000	16:11
02 *Subject Test	INT*	16:11
03 Room Air	0.000	16:12

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 22  
Expiration Date = 07/01/2012  
County = 08                          Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check      0.100 AC + acetone*

Form 106-I8000

80-004191  
p. 72 of 76  
KAL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004191  
Location = TOXL                    8164.13.00 06/09  
02/25/2011                            16:12

Test	AC	Time
01 Diagnostic	OK	16:13
02 Room Air	RFI*	16:13
03 Room Air	0.000	16:14

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 22  
Expiration Date = 07/01/2012  
County = 08                          Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*RFI test*

Form 106-I8000

80-004191  
p. 73 of 76  
YML

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = TOXL      8164.13.00 06/09  
02/25/2011      15:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:01
02 Std. Gas	0.080	15:01
03 Room Air	0.000	15:02
04 Std. Gas	0.081	15:02
05 Room Air	0.000	15:03
06 Std. Gas	0.081	15:03
07 Room Air	0.000	15:03

Lot No = 15410080A1  
Cyl No = 22  
Exp Date = 07/01/2012  
County = 08      Oper No. = 777777

*Deb Kashur*  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004191  
p. 74 of 76  
VEM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004191  
Location = TOXL      8164.13.00 06/09  
02/25/2011      15:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:23
02 Std. Gas	0.081	15:24
03 Room Air	0.000	15:24
04 Std. Gas	0.081	15:25
05 Room Air	0.000	15:25
06 Std. Gas	0.080	15:25
07 Room Air	0.000	15:26

Lot No = 15410080A1  
Cyl No = 22  
Exp Date = 07/01/2012  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-004191  
P. 75 of 76  
JKM




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004191  
Location = TOXL                 8164.13.00 06/09  
02/25/2011                        15:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:28
02 Std. Gas	0.081	15:29
03 Room Air	0.000	15:29
04 Std. Gas	0.081	15:30
05 Room Air	0.000	15:30
06 Std. Gas	0.080	15:30
07 Room Air	0.000	15:31

Lot No = 15410080A1  
Cyl No = 22  
Exp Date = 07/01/2012  
County = 08                        Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:  
*Calibration check 0.080 AC*

Form 106-I8000

80-004191  
p. 76 of 76  
JK