



ETHANOL GAS STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>HERZIG THOMAS J</i>		
Location <i>LAUREL SO</i>	Intoxilyzer® Serial Number <i>80-005956</i>	
Gas Lot Number <i>30</i>	Gas Cylinder Number <i>29614080A2</i>	Gas Expiration Date <i>12/05/2016</i>

Deb Shanaver

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.078* AC
 - B. *0.079* AC
 - C. *0.079* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>02/25/16 ✓</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Shanaver</i>	Date <i>01 Mar 2016</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005956
Location = LAMO 8164.13.00 06/09
02/25/2016 12:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:17
02 Std. Gas	0.078	12:17
03 Room Air	0.000	12:18
04 Std. Gas	0.079	12:18
05 Room Air	0.000	12:19
06 Std. Gas	0.079	12:19
07 Room Air	0.000	12:20

Lot No = 29614080A2
Cyl No = 30
Exp Date = 12/05/2016
County = 23 Oper No. = 039706


Operator Signature
THOMAS HERZIG

Remarks:

Form 106-I8000

New Form