

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Chris Pulver

Location: Mountrail SO

Intoxilyzer Serial No.: 80-003071

Lot Number: 659358

Cylinder Number: 43

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.077%

Test 2: 0.079%

Test 3: 0.079%

(ATTACH TEST RECORD)

Chris Pulver
Chemical Test Operator Signature

2-26-10
Date

Deb Kashner
Reviewed By

23 Mar 2010
Date

80-003071
p.1 of 2

7/2/09


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = STAN 8164.13.00 06/09
02/26/2010 10:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:38
02 Std. Gas	0.077	10:38
03 Room Air	0.000	10:39
04 Std. Gas	0.079	10:39
05 Room Air	0.000	10:40
06 Std. Gas	0.079	10:40
07 Room Air	0.000	10:41

Lot No = 659358
Cyl No = 43
Exp Date = 09/03/2011
County = 31 Oper No. = 049912



Operator Signature
CHRISTOPHER PULVER

Remarks:

Form 106-I8000

80-003071
P. 2 of 2 vda
7/9/09