



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59281 (5-2011)

Serial Number 80-002666	Instrument Location HARV
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input checked="" type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (Use CMS Mode; Key Radio During First Room Air).
- 7. Repair and/or Maintenance Performed (if Any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Mark Countess</i>	Date 2-27-13
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date 01 Mar 2013

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
02/27/2013 14:58

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 22
Standard Expiration? 10/01/2013
Oper No? 030819

Flow Cal. Date: 12/31/2012
Slope 657
Intercept -418473

IR Calibration Date: 12/31/2012
 3um 9um

0th Coef(*100): -31967 -16114
1st Coef(*100): 279454 136158
2nd Coef(*100): 1837 1190
H2O adj(mg/l*10k): 527 293

***** Printer Test End *****



Operator Signature
MARK COUNTESS

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
02/27/2013 14:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:48
02 Std. Gas	0.079	14:48
03 Room Air	0.000	14:48
04 Std. Gas	0.079	14:49
05 Room Air	0.000	14:49
06 Std. Gas	0.079	14:50
07 Room Air	0.000	14:50

Lot No = 23411080A1

Cyl No = 22

Exp Date = 10/01/2013

County = 52

Oper No. = 030819



Operator Signature
MARK COUNTESS

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = HARV 8164.13.00 06/09
02/27/2013 14:51

Test	AC	Time
01 Diagnostic	OK	14:53
02 Room Air	RFI*	14:53
03 Room Air	0.000	14:54

*Invalid Test
Inhibited - RFI

Sub Name = RFI, TEST NA
Sub DOB = 01/01/2013
Sub Sex = Male Weight = 123
Test = DUI Cit = 12345
Dr. Lic. = ND/RFI 123
Lot No = 23411080A1
Cyl No = 22
Expiration Date = 10/01/2013
County = 52 Oper No. = 030819

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
MARK COUNTESS

Remarks: