



**OFFICE OF ATTORNEY GENERAL**  
**CRIME LABORATORY DIVISION**  
 Toxicology Section/Breath Alcohol Program  
 Instrument Certification and Maintenance Unit

BrF-005-8

- Q. Flow Monitor. Level 3 Func D Sub F. R 181 F 0.000 L/s
- R. Temperature Monitor. Level 3 Func D Sub T  
 C 47.0 A 48.5 I 0.0  
 B 45.0 V 0.0
- S. Gas Tank Monitor. Level 3 Func D Sub G  
 T. Configure Wet Bath Sim. Level 1 Func S  
214 psi Intox 230 psi Reg
- U. Zero Alcohol Test Level 1 Func C  
DR5188 Sim SN — Lot 0.000 Rep Value  
0.000 AC 0.000 AC 0.000 AC
- V. Low AC Test. Level 1 Func C  
DR5144 Sim SN — Lot 0.021 Rep Value  
0.020 AC 0.020 AC 0.021 AC
- W. Linearity Test. Level 1 Func C  
DR3847 Sim SN — Lot 0.294 Rep Value  
0.294 AC 0.295 AC 0.296 AC
- X. Interferent Test.  
 Level 1 Func B  
 Acetone DR3845 Sim SN — Lot 0.05% Conc  
 Display  Printout  
 Methanol G10573 Sim SN — Lot 0.10% Conc  
 Display  Printout  
 IPA DR5191 Sim SN — Lot 0.10% Conc  
 Display  Printout
- Y. Ambient Test  
 Z. Invalid Sample Test  
 AA. RFI Detector Test. CMS Mode  
 BB. Configure Dry Gas Standard. Level 1 Func S  
 CC. Stability Test at 0.080AC. Level 3 Func T Sub S  
0.0815 Ave. 0.0005 Std Dev 0.63% Rel Std Dev
- DD. Maintenance/Repairs Initial inspection w/software 8/6/13.

28 Feb 2011  
 Date  
28 Feb 2011  
 Date

Deb Kashur  
 Signature  
Mary Pearson  
 Reviewed by

80-004204  
 P. 67 of 82  
 JH

Intoxilyzer Test Record and Checklist  
NDODAG Crime Lab. Div., Bismarck, ND 58501

OMI, Inc. Intoxilyzer Alcohol Analyzer  
North Dakota Model 8000 SN 80-004204  
Location = TOXL 8164.13.00 06/09  
02/28/2011 08:40

\*\*\*\*\* Printer Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMN

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNQRSTU

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMNQRSTUWX

32 characters per line  
abcdefghijklmnopqrstuwx123456  
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line  
abcdefghijklmnopqrstuwx1234567890abcdef  
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? No  
# of Print Copies? 1  
Select Std (D/W/I)? Dry  
Standard Value? 0.080  
Standard Lot #? 15410080A1  
Standard Cyl #? 22  
Standard Expiration? 07/01/2012  
Oper No? 77777

Flow Cal. Date: 07/13/2010  
Slope 750  
Intercept -815939

IR Calibration Date: 07/13/2010

	3um	9um
--	-----	-----

0th Coef(*100):	-24262	-22939
1st Coef(*100):	266889	136772
2nd Coef(*100):	1535	960
H2O adj(mg/l)*10K):	459	425

\*\*\*\*\* Test End \*\*\*\*\*

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Print test*  
Form 106-18000

80-004204  
P. 68 of 83  
FR

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer           Alcohol Analyzer  
North Dakota Model 8000       SN 80-004204  
Location = TOXL               8164.13.00 06/09  
02/28/2011                       08:40

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxy1234567890-==|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxy1234567890-==|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:           Enabled  
Start Test Sequence:       DABACABA  
Display Prelim Rslt?       Yes  
Display Third Digit?       Yes  
Inhib Printer(Y/N)?       No  
Display Volume?           No  
Disable On Memfull?       No  
# of Print Copies?         1  
Select Std (D/W/I)?       Dry  
Standard Value?           0.080  
Standard Lot #?           15410080A1  
Standard Cyl #?           22  
Standard Expiration?      07/01/2012  
Oper No?                   777777

Flow Cal. Date:           07/13/2010  
Slope                      750  
Intercept                 -815939

IR Calibration Date:      07/13/2010

	3um	9um
0th Coef(*100):	-24262	-22939
1st Coef(*100):	266889	136772
2nd Coef(*100):	1535	960
H2O adj(mg/l*10k):	459	425

\*\*\*\*\* Printer Test End \*\*\*\*\*

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

Print test

Form 106-I8000

80-004204  
P. 69 of 82  
JAH

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                8164.13.00 06/09  
02/28/2011                        08:46

Test	AC	Time
01 Room Air	0.000	08:47
02 *Subject Test	REF*	08:47
03 Room Air	0.000	08:48

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 22  
Expiration Date = 07/01/2012  
County = 08                          Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*Camera/Refusal check*

Form 106-I8000

80-004204  
P. 70483  
JH


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                    8164.13.00 06/09  
02/28/2011                            08:48

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Diagnostic Test*  
Form 106-I8000

*80-004204  
P. 714 PA  
YH*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.13.00 06/09  
02/28/2011      09:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:22
02 Std. Sol.	0.000	09:23
03 Room Air	0.000	09:24
04 Std. Sol.	0.000	09:24
05 Room Air	0.000	09:25
06 Std. Sol.	0.000	09:25
07 Room Air	0.000	09:26

08 Sim Temp = 34.0°C

Simul Ser No = 00000  
Std Sol No = 000  
County = 08

Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

Zero alcohol 0.000 AC

Form 106-I8000

80-004204  
p. 72 of 82  
JK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.13.00 06/09  
02/28/2011      09:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:27
02 Std. Sol.	0.020	09:28
03 Room Air	0.000	09:28
04 Std. Sol.	0.020	09:29
05 Room Air	0.000	09:30
06 Std. Sol.	0.021	09:30
07 Room Air	0.000	09:31

08 Sim Temp = 34.0°C

Simul Ser No = 00000  
Std Sol No = 000  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

low AC      0.021 AC

Form 106-I8000

80-004204  
p. 73 of 82  
km



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.13.00 06/09  
02/28/2011      09:32

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:33
02 Std. Sol.	0.294	09:34
03 Room Air	0.000	09:34
04 Std. Sol.	0.295	09:35
05 Room Air	0.000	09:36
06 Std. Sol.	0.296	09:36
07 Room Air	0.000	09:37

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.294 AC*

Form 106-I8000

80-004204  
p. 74 of P2  
KM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.13.00 06/09  
02/28/2011      09:47

Test	AC	Time
01 Room Air	0.000	09:47
02 *Subject Test	INT*	09:48
03 Room Air	0.000	09:48

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 22  
Expiration Date = 07/01/2012  
County = 08      Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Interferent check w/ 0.05% acetone*

Form 106-I8000

80-009204  
P. 75482  
km

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                    8164.13.00 06/09  
02/28/2011                            09:49

Test	AC	Time
01 Room Air	0.000	09:49
02 *Subject Test	INT*	09:50
03 Room Air	0.000	09:50

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 22  
Expiration Date = 07/01/2012  
County = 08                          Oper No. = 777777



Operator Signature  
DEB KASHUR

Remarks:

*Interferent check w/ 0.10% Methanol*

Form 106-I8000

80-004204  
P. 76 of 82  
jm

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004204  
Location = TOXL                8164.13.00 06/09  
02/28/2011                        09:51

Test	AC	Time
01 Room Air	0.000	09:51
02 *Subject Test	INT*	09:52
03 Room Air	0.000	09:52

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 22  
Expiration Date = 07/01/2012  
County = 08                          Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check w/0.1090 IPA*

Form 106-I8000

80-004204  
P. 77d P2  
KH

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004204  
Location = TOXL                    8164.13.00 06/09  
02/28/2011                            09:57

Test	AC	Time
01 Room Air	AMB*	09:58
02 Room Air	0.000	09:59

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 22  
Expiration Date = 07/01/2012  
County = 08                            Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*Ambient fail test*

Form 106-I8000

80-004204  
P. 78 of 82  
FA

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.13.00 06/09  
02/28/2011      10:00

Test	AC	Time
01 Room Air	0.000	10:00
02 Invalid Sample	X.XXX	10:01
03 Room Air	0.000	10:01

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 22  
Expiration Date = 07/01/2012  
County = 08      Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Invalid Sample Test*  
Form 106-I8000

80-004204  
p. 79 of 82  
FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004204  
Location = TOXL                    8164.13.00 06/09  
02/28/2011                            10:04

Test	AC	Time
01 Diagnostic	OK	10:05
02 Room Air	RFI*	10:05
03 Room Air	0.000	10:05

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 22  
Expiration Date = 07/01/2012  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

80-004204  
p. 80 of 80  
sm

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004204  
Location = TOXL      8164.13.00 06/09  
02/28/2011      10:07

Test	AC	Time
01 Room Air	0.000	10:07
02 Std. Gas	0.081	10:08
03 Room Air	0.000	10:08
04 Std. Gas	0.081	10:09
05 Room Air	0.000	10:09
06 Std. Gas	0.081	10:10
07 Room Air	0.000	10:10
08 Std. Gas	0.081	10:10
09 Room Air	0.000	10:11
10 Std. Gas	0.081	10:11
11 Room Air	0.000	10:12
12 Std. Gas	0.082	10:12
13 Room Air	0.000	10:13
14 Std. Gas	0.081	10:13
15 Room Air	0.000	10:14
16 Std. Gas	0.082	10:14
17 Room Air	0.000	10:14
18 Std. Gas	0.081	10:15
19 Room Air	0.000	10:15
20 Std. Gas	0.082	10:16
21 Room Air	0.000	10:16
22 Std. Gas	0.082	10:17
23 Room Air	0.000	10:17
24 Std. Gas	0.082	10:17
25 Room Air	0.000	10:18
26 Std. Gas	0.082	10:18
27 Room Air	0.000	10:19
28 Std. Gas	0.081	10:19
29 Room Air	0.000	10:20
30 Std. Gas	0.082	10:20
31 Room Air	0.000	10:20
32 Std. Gas	0.082	10:21
33 Room Air	0.000	10:21
34 Std. Gas	0.082	10:22
35 Room Air	0.000	10:22
36 Std. Gas	0.082	10:23
37 Room Air	0.000	10:23
38 Std. Gas	0.081	10:23
39 Room Air	0.000	10:24
40 Std. Gas	0.081	10:24
41 Room Air	0.000	10:25
42 Cal Check Stats:		
43 Average	0.0815	
44 Std Dev	0.0005	
45 Rel Std Dev	0.63%	

*Stability test  
Deb Keshur  
p. 1 of 2*

*80-004204  
p. 81 of 82  
FM*



Operator Signature  
DEB KASHUR

Remarks:

Lot #: 15410080A1 Cyl #: 22 EXP: 07/01/2012

Form 106-I8000

Stability test  
Deb Kashur  
p. 2 of 2

80-004204  
P. 82 of 82  
FM