

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004204

Location: TOXL

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5144</u> Lot No. <u> </u> Rep AC <u>0.021</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u> </u> Rep AC <u>0.294</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3452</u> Lot No. <u> </u> Rep AC <u>0.100</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>15410080A1</u>	Cyl No. <u>22</u>	Exp. Date <u>07/01/2012</u>
Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.082</u>	Test 5 <u>0.082</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.082</u>	Test 9 <u>0.082</u>
Average <u>0.081</u>		

L. Tank pressure, Level 3 Func D Sub G P: 133 psi Reg 130 psi

M. Remarks/Maintenance record: Annual inspection w/software 8/64.13

28 Feb 2011
Date

Deb Kashur
Field Inspector's Signature

28 Feb 2011
Date

Mary Beam
Reviewed by

80-004204
p. 83 of 91
Jm

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
02/28/2011 10:58

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 15410080A1
Standard Cyl #? 22
Standard Expiration? 07/01/2012
Oper No? 777777

Flow Cal. Date: 07/13/2010
Slope 750
Intercept -815939

IR Calibration Date: 07/13/2010
 3um 9um

0th Coef(*100): -24262 -22939
1st Coef(*100): 266889 136772
2nd Coef(*100): 1535 960
H2O adj(mg/l*10k): 459 425

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:

Print Test

Form 106-I8000

80-004204
p. 84 of 91
KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
02/28/2011 10:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:00
02 Std. Sol.	0.021	11:00
03 Room Air	0.000	11:01
04 Std. Sol.	0.021	11:02
05 Room Air	0.000	11:02
06 Std. Sol.	0.021	11:03
07 Room Air	0.000	11:03

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.021AC

Form 106-I8000

*80-004204
p. 85 of 91
km*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
02/28/2011 11:04

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:05
02 Std. Sol.	0.294	11:09
03 Room Air	0.000	11:09
04 Std. Sol.	0.296	11:10
05 Room Air	0.000	11:10
06 Std. Sol.	0.297	11:11
07 Room Air	0.000	11:12

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

~~Print test~~ ^{DK} Linearity test 0.294 AC

Form 106-I8000

80-004204
P-86 of 91
JEM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
02/28/2011 11:14

Test	AC	Time
01 Room Air	0.000	11:15
02 *Subject Test	INT*	11:15
03 Room Air	0.000	11:16

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 1541008A1
Cyl No = 22
Expiration Date = 07/01/2012
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Interferent test 0.100 AC + acetone

Form 106-I8000

80-004204
p. 87 of 91
JKM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
02/28/2011 11:17

Test	AC	Time
01 Diagnostic	OK	11:18
02 Room Air	RFI*	11:18
03 Room Air	0.000	11:18

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 1541008A1
Cyl No = 22
Expiration Date = 07/01/2012
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

*80-004204
p. 88-191
km*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
02/28/2011 11:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:20
02 Std. Gas	0.081	11:20
03 Room Air	0.000	11:20
04 Std. Gas	0.082	11:21
05 Room Air	0.000	11:21
06 Std. Gas	0.081	11:22
07 Room Air	0.000	11:22

Lot No = 1541008A1
Cyl No = 22
Exp Date = 07/01/2012
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Form 106-I8000

80-004204
P. 89 of 91
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
02/28/2011 11:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:23
02 Std. Gas	0.081	11:24
03 Room Air	0.000	11:24
04 Std. Gas	0.082	11:25
05 Room Air	0.000	11:25
06 Std. Gas	0.082	11:25
07 Room Air	0.000	11:26

Lot No = 1541008A1
Cyl No = 22
Exp Date = 07/01/2012
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

*80-004204
p. 90 of 91
FK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
02/28/2011 11:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:29
02 Std. Gas	0.081	11:29
03 Room Air	0.000	11:29
04 Std. Gas	0.081	11:30
05 Room Air	0.000	11:30
06 Std. Gas	0.082	11:31
07 Room Air	0.000	11:31

Lot No = 1541008A1
Cyl No = 22
Exp Date = 07/01/2012
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004204
p. 91 of 91
FM