

INTOX
PBT
TOXITEST

DATE 3-1-10
RECEIVED VIA Fidup

CMI INC.

316 E. 9th St., Owensboro, KY 42303
270-685-6200 — Fax 270-685-6288
1-800-835-0690

SERVICE WORK ORDER

4323175

B I L L

ATTN:

CITY STATE ZIP

CUSTOMER NO.
585013

PHONE

S H I P

ATTN:
North Dakota Crime Lab
2641 East Main Avenue
CITY STATE ZIP
Bismarck, ND 58501

PHONE

MODEL # 8000	S/N 80-04947	DATE RCVD.	P.O. #	REQ.	BILL CODE 3	SERVICE LOCATION CODE
PROD. CODE	EXTRA PARTS RCVD.					

DESCRIPTION OF PROBLEM

"Won't come out of diagnostics - Int Printer Fehl."

QUOTE WARRANTY
 SERVICE DEPT. NEW

PRELIMINARY TEST DISMANTLE, INSPECT, DIAGNOSTIC TEST

POWER SUPPLY TEST, INSPECT REPLACE

MAIN PWB's - REMOVE & CLEAN, REINSTALL, TEST

PWB's DIAGNOSE, REPAIR REPLACE

PWB's - SURFACE CLEAN IN PLACE

FILTER WHEEL - REMOVE, CLEAN, REINSTALL, CHECK CELL INTERIOR

CHOPPER WHEEL MOTOR TEST & INSPECT REPLACE

CELL - REMOVE, CLEAN, REINSTALL, REPLACE LENSES/GASKETS REQ'D

IR LIGHT SOURCE REMOVE, REALIGN REPLACE

PRINTER ASSY. DISASSEMBLE CLEAN & LUBRICATE

REMOVE & REPLACE BELT REALIGNMENT

REPLACE TEST PRINTER FUNCTIONS

BATTERIES - REPLACE CPU RAM

REPLACE PARTS REASSEMBLE CALIBRATION FINAL TEST

MISC. _____

PBT PRELIMINARY TESTING - PRESSURE, BATTERY, ZERO SET

DISASSEMBLE REPLACE PARTS REASSEMBLE

TEST CALIBRATE FINAL TEST

TOXITEST INCOMING LEAK TEST VISUAL/MECH. INSPECTION

TEST THERMOMETER & THERMOSTAT DISASSEMBLE

REPLACE PARTS REASSEMBLE OUTGOING & LEAK TEST

FINAL TEST

MISC. _____

PARTS USED

QTY	ID NO.	DESCRIPTION	CODE

COMMENTS:

Instrument operating within specifications. I had the unit re-fine tested.

135

HOURS: 1 hour

TECH: [Signature]

DATE: 3-1-10

CALL TECH. DATE

ID: # 53176

SHIPPING DATE/CHARGES

80-004947

P. 32/33 15m

SPECIAL SHIPPING INSTRUCTIONS

Statement of Warranty

New Product Warranty

CMI Inc. warrants that each new product will be free from defects in material and workmanship, under normal use and service, for a period of one year from the date of invoice to the initial purchaser. CMI's obligation is limited to repairing or replacing, as CMI may elect, any part or parts of such product, which CMI determines to be defective in material or workmanship. Warranty repairs will be performed at the factory or at a factory authorized service center.

The product, or part of the product, considered to be covered by the conditions of this warranty shall be returned, freight prepaid, in its original shipping container or similar protective container, to the factory, only after receipt of a Returned Material Authorization number from CMI. The repaired or replacement part or product will be returned from CMI or the authorized service center, freight prepaid.

Warranty coverage extends only to the original purchaser and does not include abuse, misuse, cables, switches or use of the product for other than its intended purpose. This warranty also does not apply if the product is adversely affected by attaching any feature or device to it, or is in any way tampered with or modified, without expressed written permission from CMI, Inc.

Repaired Product Warranty

Out of warranty product repairs are warranted for 90 days from the date of repair. This includes labor and those parts, which are replaced. If additional repair is required within the 90-day periods, there will be a charge for any parts that were not originally replaced. Repairs made during this 90-day period that are unrelated to the original repair are not covered under the warranty.

There are no warranties expressed or implied, including but not limited to, other than those contained in this warranty. In no event shall CMI be liable for any loss of profits or any indirect or consequential damages arising out of any such defect in material or workmanship.

INTOXILYZER[®]
— **CMI** —

316 E. 9th Street
Owensboro, KY 42303
1-800-835-0690
www.alcoholtest.com

80-004987
P. 33/33