

CMI INC.

316 E. 9th St., Owensboro, KY 42303
 270-685-6200 — Fax 270-685-6288
 1-800-835-0690

SERVICE WORK ORDER

#323076

INTOX
PBT
TOXITEST

DATE 3-1-10
 RECEIVED VIA Feb 28

B
I
L
L
T
O

ATTN:
 CITY STATE ZIP
 CUSTOMER NO. 585013
 PHONE

S
H
I
P
T
O

ATTN:
 N.D. Crime Laboratory
 2641 East Main Avenue
 CITY STATE ZIP
 Bismarck ND 58501
 PHONE

MODEL #	S/N	DATE RCVD.	P.O. #	REQ.	BILL CODE	SERVICE LOCATION CODE
8000	80-004952				3	
PROD. CODE	EXTRA PARTS RCVD.					
	(3) Gauss Cylinder - SWO#					

DESCRIPTION OF PROBLEM: "Screw for battery pack will not inscrew final case."
 QUOTE WARRANTY SERVICE DEPT. NEW

- PRELIMINARY TEST DISMANTLE, INSPECT, DIAGNOSTIC TEST
- POWER SUPPLY TEST, INSPECT REPLACE
- MAIN PWB's - REMOVE & CLEAN, REINSTALL, TEST
- PWB's DIAGNOSE, REPAIR REPLACE
- PWB's - SURFACE CLEAN IN PLACE
- FILTER WHEEL - REMOVE, CLEAN, REINSTALL, CHECK CELL INTERIOR
- CHOPPER WHEEL MOTOR TEST & INSPECT REPLACE
- CELL - REMOVE, CLEAN, REINSTALL, REPLACE LENSES/GASKETS REQ'D
- IR LIGHT SOURCE REMOVE, REALIGN REPLACE
- PRINTER ASSY. DISASSEMBLE CLEAN & LUBRICATE
- REMOVE & REPLACE BELT REALIGNMENT
- REPLACE TEST PRINTER FUNCTIONS
- BATTERIES - REPLACE CPU RAM
- REPLACE PARTS REASSEMBLE CALIBRATION FINAL TEST
- MISC.
- PBT PRELIMINARY TESTING - PRESSURE, BATTERY, ZERO SET
- DISASSEMBLE REPLACE PARTS REASSEMBLE
- TEST CALIBRATE FINAL TEST
- TOXITEST INCOMING LEAK TEST VISUAL/MECH. INSPECTION
- TEST THERMOMETER & THERMOSTAT DISASSEMBLE
- REPLACE PARTS REASSEMBLE OUTGOING & LEAK TEST
- FINAL TEST
- MISC.

PARTS USED			
QTY	ID NO.	DESCRIPTION	CODE
1	416278	Insert for plastic, M2	

COMMENTS: I verified the problem. I replaced the brass insert, I had the unit re-final tested.
 138

HOURS: 1 hour ID: # 53174
 TECH: *[Signature]* DATE: 3-1-10 CAL TECH: _____ DATE: _____

SHIPPING DATE/CHARGES: 80-004952 P. 4/1 AM
 SPECIAL SHIPPING INSTRUCTIONS:

Statement of Warranty

New Product Warranty

CMI Inc. warrants that each new product will be free from defects in material and workmanship, under normal use and service, for a period of one year from the date of invoice to the initial purchaser. CMI's obligation is limited to repairing or replacing, as CMI may elect, any part or parts of such product, which CMI determines to be defective in material or workmanship. Warranty repairs will be performed at the factory or at a factory authorized service center.

The product, or part of the product, considered to be covered by the conditions of this warranty shall be returned, freight prepaid, in its original shipping container or similar protective container, to the factory, only after receipt of a Returned Material Authorization number from CMI. The repaired or replacement part or product will be returned from CMI or the authorized service center, freight prepaid.

Warranty coverage extends only to the original purchaser and does not include abuse, misuse, cables, switches or use of the product for other than its intended purpose. This warranty also does not apply if the product is adversely affected by attaching any feature or device to it, or is in any way tampered with or modified, without expressed written permission from CMI, Inc.

Repaired Product Warranty

Out of warranty product repairs are warranted for 90 days from the date of repair. This includes labor and those parts, which are replaced. If additional repair is required within the 90-day periods, there will be a charge for any parts that were not originally replaced. Repairs made during this 90-day period that are unrelated to the original repair are not covered under the warranty.

There are no warranties expressed or implied, including but not limited to, other than those contained in this warranty. In no event shall CMI be liable for any loss of profits or any indirect or consequential damages arising out of any such defect in material or workmanship.

INTOXILYZER®
— **CMI** —

316 E. 9th Street
Owensboro, KY 42303
1-800-835-0690
www.alcoholtest.com

80-004952
P. 7/7 VML