



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David M Wallace</u>		
Location: <u>Dickinson LLC</u>		
Intoxilyzer® Serial Number: <u>80-004936</u>		
Gas Lot Number: <u>15410080R1</u>	Gas Cylinder Number: <u>20</u>	Expiration Date: <u>7-1-12</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.029</u> %		
Test 2: <u>0.029</u> %		
Test 3: <u>0.028</u> %		
(ATTACH TEST RECORD)		

David M Wallace
Chemical Test Operator Signature

3-1-11
Date

Deb Kashur
Reviewed By

03 Mar 2011
Date

80-004936
p. 70 of 71
rm

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = DKPD 8164.13.00 06/09
03/01/2011 08:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:16
02 Std. Gas	0.079	08:17
03 Room Air	0.000	08:17
04 Std. Gas	0.079	08:18
05 Room Air	0.000	08:18
06 Std. Gas	0.078	08:18
07 Room Air	0.000	08:19

Lot No = 15410080A1
Cyl No = 20
Exp Date = 07/01/2012
County = 45 Oper No. = 068332



Operator Signature
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-004936
P. 71 of 71
FN