



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: SHANNON P. DUISSEN		
Location: TIOGA		
Intoxilyzer® Serial Number: 80-003071		
Gas Lot Number: 15410080A1	Gas Cylinder Number: 25	Expiration Date: 7/1/12
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0.078 %		
Test 2: 0.079 %		
Test 3: 0.078 %		
(ATTACH TEST RECORD)		

Shannon P. Duisen
Chemical Test Operator Signature

3/2/11
Date

Deb Kashner
Reviewed By

23 Mar 2011
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TIOG 8164.13.00 06/09
03/02/2011 10:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:47
02 Std. Gas	0.078	10:48
03 Room Air	0.000	10:48
04 Std. Gas	0.079	10:48
05 Room Air	0.000	10:49
06 Std. Gas	0.078	10:49
07 Room Air	0.000	10:50

Lot No = 15410080A1
Cyl No = 25
Exp Date = 07/01/2012
County = 53 Oper No. = 020632



Operator Signature
SHANNON P. DUISEN

Remarks:

Form 106-I8000

80-003071
P. 117 of 117
RM