



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59281 (5-2011)

Serial Number 80 - 005452	Instrument Location Burleigh Co. Sheriff's Dept.
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (Use CMS Mode; Key Radio During First Room Air).
- 7. Repair and/or Maintenance Performed (if Any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date 3-3-2014 ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date 07 Mar 2014

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: *Charles E. Eda* Certified Date: **21 Nov. 16**

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BCSO 8164.13.00 06/09
03/03/2014 10:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 69
Standard Expiration? 08/01/2014
Oper No? 130100

Flow Cal. Date: 06/21/2013
Slope 734
Intercept -865305

IR Calibration Date: 06/21/2013

	3um	9um
0th Coef(*100):	-21281	-23057
1st Coef(*100):	292132	137092
2nd Coef(*100):	2334	1231
H2O adj(mg/l*10k):	497	482

***** Printer Test End *****



Operator Signature
JEFFREY VYSKA

Remarks:

Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BCSO 8164.13.00 06/09
03/03/2014 10:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:32
02 Std. Gas	0.081	10:32
03 Room Air	0.000	10:33
04 Std. Gas	0.081	10:33
05 Room Air	0.000	10:34
06 Std. Gas	0.081	10:34
07 Room Air	0.000	10:35

Lot No = 16512080A2
Cyl No = 69
Exp Date = 08/01/2014
County = 08 Oper No. = 130100


Operator Signature
JEFFREY VYSKA

Remarks:

ACA Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005952
Location = BCSO 8164.13.00 06/09
03/03/2014 10:36

Test	AC	Time
01 Diagnostic	OK	10:39
02 Room Air	0.000	10:40
03 *Subject Test	RFI*	10:40
04 Room Air	0.000	10:40

*Invalid Test
Inhibited - RFI

Sub Name = RFI, NA NA
Sub DOB = 08/03/1979
Sub Sex = Male Weight = 210
Test = DUI Cit = 0
Dr. Lic. = ND/0
Lot No = 16512080A2
Cyl No = 69
Expiration Date = 08/01/2014
County = 08 Oper No. = 130100

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.


Operator Signature
JEFFREY VYSKA

Remarks:

RFI Test

Form 106-I8000