



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Roy Weber</u>		
Location: <u>TRAIL CO S.O.</u>		
Intoxilyzer® Serial Number: <u>80-004200</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>#45</u>	Expiration Date: <u>09/03/2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>081</u> % Test 2: 0. <u>081</u> % Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

Roy Weber
Chemical Test Operator Signature

03/05/10
Date

Deb Kashner
Reviewed By

12 May 2010
Date

80-004200
p. 72/3 1st


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = HILL 8164.13.00 06/09
03/05/2010 21:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:56
02 Std. Gas	0.081	21:56
03 Room Air	0.000	21:57
04 Std. Gas	0.081	21:57
05 Room Air	0.000	21:58
06 Std. Gas	0.081	21:58
07 Room Air	0.000	21:59

Lot No = 659358
Cyl No = 45
Exp Date = 09/03/2011
County = 49 Oper No. = 058405



Operator Signature
RAY G. WEBER

Remarks:

Form 106-I8000

80-004200
P. 73/73 FR