



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Hinton, Rick W</u>		
Location: <u>Cass County Jail</u>		
Intoxilyzer @ Serial Number: <u>80-004946</u>		
Lot Number: <u>15410080A1</u>	Cylinder Number: <u>10</u>	Expiration Date: <u>7-1-12</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.077</u> % <i>JK</i>		
Test 2: <u>0.077</u> %		
Test 3: <u>0.078</u> %		
(ATTACH TEST RECORD)		

Rick W Hinton
Chemical Test Operator Signature

3-6-11
Date

Deb Kashner
Reviewed By

23 Mar 2011
Date

80-004946
p. 71 of 72
SM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = CASS 8164.13.00 06/09
03/06/2011 04:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	05:00
02 Std. Gas	0.077	05:00
03 Room Air	0.000	05:01
04 Std. Gas	0.077	05:01
05 Room Air	0.000	05:02
06 Std. Gas	0.078	05:02
07 Room Air	0.000	05:03

DK per Dan Herman
Lot No = ~~668180~~ 15410080A1
Cyl No = ~~76~~ 10
Exp Date = ~~09/21/2011~~ 7-1-12
County = 09 Oper No. = 029920



Operator Signature
RICK W. HINTON

Remarks:

Form 106-I8000

80-004946
P. 72 of 72
FM