

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David M. Walker

Location: Bill Kinison & Co

Intoxilyzer Serial No.: 80-003066

Lot Number: 45-9358

Cylinder Number: 23

Expiration Date: 9-3-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.082%

Test 2: 0.082%

Test 3: 0.082%

(ATTACH TEST RECORD)

David M. Walker
Chemical Test Operator Signature

3-8-10
Date

Deb Kashner
Reviewed By

10 Mar 2010
Date

80-003066
P.1 of 2 gm


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = DKPD 8164.13.00 06/09
03/08/2010 09:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:38
02 Std. Gas	0.082	09:38
03 Room Air	0.000	09:39
04 Std. Gas	0.082	09:39
05 Room Air	0.000	09:40
06 Std. Gas	0.082	09:40
07 Room Air	0.000	09:41

Lot No = 659358
Cyl No = 23
Exp Date = 09/03/2011
County = 45 Oper No. = 068332


Operator Signature
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-003066
P. 2 of 2
sam