



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Jamie M Huschka</i>		
Location <i>Williams Co LEC</i>	Intoxilyzer® Serial Number <i>80-004950</i>	
Gas Lot Number <i>23411080A1</i>	Gas Cylinder Number <i>50</i>	Gas Expiration Date <i>10/01/2013</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.079* AC
 - B. *0.079* AC
 - C. *0.080* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Jamie M Huschka</i>	Date <i>08 March 2013</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kasher (Rec'd 18 Feb 14)</i>	Date <i>18 Feb 14</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004950
Location = WILL 8164.13.00 06/09
03/08/2013 20:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:41
02 Std. Gas	0.079	20:41
03 Room Air	0.000	20:42
04 Std. Gas	0.079	20:42
05 Room Air	0.000	20:43
06 Std. Gas	0.080	20:43
07 Room Air	0.000	20:44

Lot No = 23411080A1
Cyl No = 50
Exp Date = 10/01/2013
County = 53 Oper No. = 100406


Operator Signature
JAMIE HUSCHKA

Remarks: