

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-004940

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5144 Lot No. Rep AC 0.019 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5187 Lot No. Rep AC 0.285 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR5114 Lot No. Rep AC 0.100 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 709352 Cyl No. 22 Exp. Date 23 Dec 2011

Test 1 0.080 Test 4 0.080 Test 7 0.079

Test 2 0.078 Test 5 0.079 Test 8 0.079

Test 3 0.080 Test 6 0.079 Test 9 0.079

Average 0.079

L. Remarks/Maintenance record: Annual inspection

11 Mar 2010
Date

Deb Kashner
Field Inspector's Signature

15 Mar 2010
Date

Mary Pearson
Reviewed by

80-004940
P.23/31 FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
03/11/2010 10:43

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 709352
Standard Cyl #? 22
Standard Expiration? 12/23/2011
Oper No? 777777

Flow Cal. Date: 01/18/2010
Slope 720
Intercept -746294

IR Calibration Date: 01/18/2010


	3um	9um
0th Coef(*100):	-25087	-26311
1st Coef(*100):	250220	132895
2nd Coef(*100):	3837	1322
H2O adj(mg/l*10k):	574	434

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:



Form 106-I8000

80-004940
2/31/10

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
03/11/2010 14:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:19
02 Std. Sol.	0.018	14:20
03 Room Air	0.000	14:20
04 Std. Sol.	0.018	14:21
05 Room Air	0.000	14:21
06 Std. Sol.	0.018	14:22
07 Room Air	0.000	14:23

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.019 AC

Form 106-18000

80-004940
25/31 YH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
03/11/2010 12:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:07
02 Std. Sol.	0.291	12:07
03 Room Air	0.000	12:08
04 Std. Sol.	0.291	12:09
05 Room Air	0.000	12:09
06 Std. Sol.	0.291	12:10
07 Room Air	0.000	12:10

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.285AC

Form 106-18000

80-004940
20/31 KFL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
03/11/2010 14:31

Test	AC	Time
01 Room Air	0.000	14:32
02 *Subject Test	INT*	14:32
03 Room Air	0.000	14:33

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 709352
Cyl No = 22
Expiration Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check

Form 106-I8000

80-004940
2/31 KA

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
03/11/2010 14:45

Test	AC	Time
01 Diagnostic	OK	14:45
02 Room Air	RFI*	14:45
03 Room Air	0.000	14:46

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 709352
Cyl No = 22
Expiration Date = 12/23/2011
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-18000

*80-004940
28/31 Klu*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
03/11/2010 14:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:49
02 Std. Gas	0.080	14:50
03 Room Air	0.000	14:50
04 Std. Gas	0.078	14:50
05 Room Air	0.000	14:51
06 Std. Gas	0.080	14:51
07 Room Air	0.000	14:52

Lot No = 709352
Cyl No = 22
Exp Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-18000

*80-004940
09/31/11*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
03/11/2010 14:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:53
02 Std. Gas	0.080	14:53
03 Room Air	0.000	14:54
04 Std. Gas	0.079	14:54
05 Room Air	0.000	14:55
06 Std. Gas	0.079	14:55
07 Room Air	0.000	14:56

Lot No = 709352
Cyl No = 22
Exp Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC
Form 106-I8000

*80-004940
30/31 KM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004940
Location = TOXL 8164.13.00 06/09
03/11/2010 14:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:56
02 Std. Gas	0.079	14:57
03 Room Air	0.000	14:57
04 Std. Gas	0.079	14:58
05 Room Air	0.000	14:58
06 Std. Gas	0.079	14:58
07 Room Air	0.000	14:59

Lot No = 709352
Cyl No = 22
Exp Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks: ~~0.079~~ ^{OK} Calibration check 0.0807AC

Form 106-I8000

80-004940
3/31/11