



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Pete Nielsen</u>		
Location: <u>West Fargo PD</u>		
Intoxilyzer® Serial Number: <u>80-003059</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>10</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>081</u> %		
Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

[Signature]  
Chemical Test Operator Signature

3-11-11  
Date

Deb Kashun  
Reviewed By

22 Mar 2011  
Date

80-003059  
P. 103 of 104  
[Signature]


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003059  
Location = WFPD                    8164.13.00 06/09  
03/11/2011                            14:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:13
02 Std. Gas	0.080	14:13
03 Room Air	0.000	14:13
04 Std. Gas	0.081	14:14
05 Room Air	0.000	14:14
06 Std. Gas	0.081	14:15
07 Room Air	0.000	14:15

Lot No = 659357  
Cyl No = 10  
Exp Date = 09/03/2011  
County = 09                            Oper No. = 039617

  
\_\_\_\_\_  
Operator Signature  
PETER NIELSEN

Remarks:

Form 106-I8000

80-00 3059  
P. 104 of 104  
VPM