

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-004934

Location: TAXI

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5144 Lot No. Rep AC 0.019 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5187 Lot No. Rep AC 0.285 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR5114 Lot No. Rep AC 0.100 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 709352 Cyl No. 22 Exp. Date 23 Dec 2011

Test 1 0.079 Test 4 0.079 Test 7 0.078

Test 2 0.079 Test 5 0.079 Test 8 0.078

Test 3 0.080 Test 6 0.079 Test 9 0.078

Average 0.079

L. Remarks/Maintenance record: Annual inspection

12 Mar 2010
Date

Deb Kashur
Field Inspector's Signature

15 Mar 2010
Date

Mary Ryan
Reviewed by

80-004934
24/32 FR

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
03/12/2010 13:30

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNQPQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNQPQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 709352
Standard Cyl #? 22
Standard Expiration? 12/23/2011
Oper No? 777777

Flow Cal. Date: 01/20/2010
Slope 698
Intercept -694859

IR Calibration Date: 03/12/2010
 3um 9um

0th Coef(*100): -18922 -21325
1st Coef(*100): 252748 128359
2nd Coef(*100): 509 713
H2O adj(mg/l*10k): 732 642

***** Printer Test End *****

 Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
03/12/2010 14:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:22
02 Std. Sol.	0.020	14:23
03 Room Air	0.000	14:23
04 Std. Sol.	0.020	14:24
05 Room Air	0.000	14:24
06 Std. Sol.	0.020	14:25
07 Room Air	0.000	14:26

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.019AC

Form 106-I8000

80-004934
26/32 *AK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
03/12/2010 14:10

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:11
02 Std. Sol.	0.286	14:12
03 Room Air	0.000	14:12
04 Std. Sol.	0.287	14:13
05 Room Air	0.000	14:13
06 Std. Sol.	0.286	14:14
07 Room Air	0.000	14:15

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.285AC

Form 106-18000

80-004934
27/32 KH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
03/12/2010 14:34

Test	AC	Time
01 Room Air	0.000	14:34
02 *Subject Test	INT*	14:35
03 Room Air	0.000	14:35

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 709352
Cyl No = 22
Expiration Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check

Form 106-I8000

80-004934
28/32 ML

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
03/12/2010 15:33

Test	AC	Time
01 Diagnostic	OK	15:34
02 Room Air	RFI*	15:34
03 Room Air	0.000	15:35

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 709352
Cyl No = 22
Expiration Date = 12/23/2011
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

80-004934
29/32 FR

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
03/12/2010 13:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:46
02 Std. Gas	0.079	13:47
03 Room Air	0.000	13:47
04 Std. Gas	0.079	13:47
05 Room Air	0.000	13:48
06 Std. Gas	0.080	13:48
07 Room Air	0.000	13:49

Lot No = 709352

Cyl No = 22

Exp Date = 12/23/2011

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-18000

80-004934
30/32 FK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
03/12/2010 15:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:21
02 Std. Gas	0.079	15:21
03 Room Air	0.000	15:22
04 Std. Gas	0.079	15:22
05 Room Air	0.000	15:23
06 Std. Gas	0.079	15:23
07 Room Air	0.000	15:23

Lot No = 709352
Cyl No = 22
Exp Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004934
3/32 KKL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
03/12/2010 15:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:36
02 Std. Gas	0.078	15:36
03 Room Air	0.000	15:36
04 Std. Gas	0.078	15:37
05 Room Air	0.000	15:37
06 Std. Gas	0.078	15:38
07 Room Air	0.000	15:38

Lot No = 709352
Cyl No = 22
Exp Date = 12/23/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-18000

80-004934
3/2/12 KH