

04/19/2010 16:12 NDOAG - Crime Laboratory Div

(FAX)701 328 6185

P.002/003



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

| | | |
|--|--------------------------------|----------------------------------|
| Chemical Test Operator Name: <u>Mark McDonald</u> | | |
| Location: <u>Valley City Police Dept</u> | | |
| Intoxilyzer® Serial Number: <u>80-004201</u> | | |
| Gas Lot Number: <u>6593526</u> | Gas Cylinder Number: <u>24</u> | Expiration Date: <u>9/3/2011</u> |
| <input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed | | |
| Instrument Test Results (Report to 3 Digits; Example, 0.081%) | | |
| Test 1: <u>0.079</u> % | | |
| Test 2: <u>0.078</u> % | | |
| Test 3: <u>0.079</u> % | | |
| (ATTACH TEST RECORD) | | |

[Signature]
Chemical Test Operator Signature

03/12/10
Date

Deb Kashner
Reviewed By

10 May 2010
Date

80-004201
P. 74/75 192


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004201
Location = VALL 8164.13.00 06/09
03/12/2010 10:59

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:59 |
| 02 Std. Gas | 0.078 | 11:00 |
| 03 Room Air | 0.000 | 11:00 |
| 04 Std. Gas | 0.078 | 11:01 |
| 05 Room Air | 0.000 | 11:01 |
| 06 Std. Gas | 0.079 | 11:01 |
| 07 Room Air | 0.000 | 11:02 |

Lot No = 659358
Cyl No = 24
Exp Date = 09/03/2011
County = 02 Oper No. = 119502



Operator Signature
MARK W. MCDONALD

Remarks:

Form 106-I8000

80-004201
P. 75/75 KM