

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-004945

Location: TDX LAB

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5190</u> Lot No. <u>—</u> Rep AC <u>0.019</u>  | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5113</u> Lot No. <u>—</u> Rep AC <u>0.285</u>                               | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5135</u> Lot No. <u>—</u> Rep AC <u>0.10</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. 709352 Cyl No. 32 Exp. Date 23 Dec 2011

Test 1 0.080 Test 4 0.080 Test 7 0.081

Test 2 0.081 Test 5 0.081 Test 8 0.081

Test 3 0.081 Test 6 0.081 Test 9 0.081

Average 0.081 AC

L. Remarks/Maintenance record: Annual Inspection

12 Mar 2010  
Date

Margy Pearson  
Field Inspector's Signature

15 Mar 2010  
Date

Deb Kashur  
Reviewed by

80-004945  
23/31 FM



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004945  
Location = TOXL      8164.13.00 06/09  
03/12/2010      13:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:43
02 Std. Sol.	0.016	13:44
03 Room Air	0.000	13:45
04 Std. Sol.	0.015	13:45
05 Room Air	0.000	13:46
06 Std. Sol.	0.016	13:47
07 Room Air	0.000	13:47

08 Sim Temp = 34.0°C

Simul Ser No = 0000000  
Std Sol No = 00000  
County = 08      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Low AC 0.019 AC*

Form 106-I8000

*80-004945  
25/31 FM*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004945  
Location = TOXL      8164.13.00 06/09  
03/12/2010      13:54

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:54
02 Std. Sol.	0.284	14:01
03 Room Air	0.000	14:01
04 Std. Sol.	0.287	14:02
05 Room Air	0.000	14:02
06 Std. Sol.	0.288	14:03
07 Room Air	0.000	14:04

08 Sim Temp = 34.0°C

Simul Ser No = 0000000

Std Sol No = 00000

County = 08

Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Linearity 0.285 AC*

Form 106-18000

80-004945  
26/31 FH


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004945  
Location = TOXL      8164.13.00 06/09  
03/12/2010      15:18

Test	AC	Time
01 Room Air	0.000	15:18
02 *Subject Test	INT*	15:19
03 Room Air	0.000	15:19

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 11/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS821456  
Lot No = 709352  
Cyl No = 32  
Expiration Date = 12/23/2011  
County = 08      Oper No. = 107501



Operator Signature  
MARGY PEARSON

Remarks:

*Alc + Acetone Interfer*

Form 106-I8000

80-004945  
2/31/12


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004945  
Location = TOXL                    8164.13.00 06/09  
03/12/2010                            14:44

Test	AC	Time
01 Room Air	RFI*	14:45
02 Room Air	0.000	14:45

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS821456  
Lot No = 709352  
Cyl No = 32  
Expiration Date = 12/23/2011  
County = 08                          Oper No. = 107501



Operator Signature  
MARGY PEARSON

Remarks:

RFI

Form 106-I8000

80-004945  
28/31 192


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004945  
Location = TOXL      8164.13.00 06/09  
03/12/2010      14:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:20
02 Std. Gas	0.080	14:21
03 Room Air	0.000	14:21
04 Std. Gas	0.081	14:21
05 Room Air	0.000	14:22
06 Std. Gas	0.081	14:22
07 Room Air	0.000	14:23

Lot No = 709352  
Cyl No = 32  
Exp Date = 12/23/2011  
County = 08      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Cal Check ACA*

Form 106-I8000

*80-004945  
29/31 VLL*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004945  
Location = TOXL      8164.13.00 06/09  
03/12/2010      14:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:26
02 Std. Gas	0.080	14:27
03 Room Air	0.000	14:27
04 Std. Gas	0.081	14:27
05 Room Air	0.000	14:28
06 Std. Gas	0.081	14:28
07 Room Air	0.000	14:29

Lot No = 709352  
Cyl No = 32  
Exp Date = 12/23/2011  
County = 08      Oper No. = 107501

  
\_\_\_\_\_  
Operator Signature  
MARGY PEARSON

Remarks:

*Cal Check ACA*  
Form 106-I8000

80004945  
3/3, 1st




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004945  
Location = TOXL      8164.13.00 06/09  
03/12/2010      14:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:30
02 Std. Gas	0.081	14:30
03 Room Air	0.000	14:31
04 Std. Gas	0.081	14:31
05 Room Air	0.000	14:31
06 Std. Gas	0.081	14:32
07 Room Air	0.000	14:32

Lot No = 709352  
Cyl No = 32  
Exp Date = 12/23/2011  
County = 08      Oper No. = 107501

  
Operator Signature  
MARGY PEARSON

Remarks:

*ACA Cal Check*

Form 106-I8000

80-004945  
3/31 PM