

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-004954

Location: Tox Lab

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5190 Lot No. — Rep AC 0.019 AC ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5113 Lot No. — Rep AC 0.285 AC ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR5135 Lot No. — Rep AC 0.10 AC + Acetone ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 709352 Cyl No. 1 Exp. Date 23 Dec 2011

Test 1 0.081 Test 4 0.081 Test 7 0.081

Test 2 0.080 Test 5 0.081 Test 8 0.081

Test 3 0.081 Test 6 0.080 Test 9 0.081

Average 0.081

L. Remarks/Maintenance record: Annual Inspection

12 Mar 10
Date

Mary Pearson
Field Inspector's Signature

15 Mar '10
Date

Deb Kashur
Reviewed by

80-004954
23/31 KH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
03/10/2010 10:06

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNQPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNQPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 709352
Standard Cyl #? 16
Standard Expiration? 12/23/2011
Oper No? 107501

Flow Cal. Date: 01/19/2010
Slope 679
Intercept -671831

IR Calibration Date: 01/19/2010
3um 9um

0th Coef(*100): -34134 -33955
1st Coef(*100): 257873 132608
2nd Coef(*100): 1059 850
H2O adj(mg/l*10k): 762 516

***** Printer Test End *****



Operator Signature
MARGY PEARSON

Remarks:

Print

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
03/10/2010 11:58

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:58
02 Std. Sol.	0.016	11:59
03 Room Air	0.000	12:00
04 Std. Sol.	0.018	12:01
05 Room Air	0.000	12:01
06 Std. Sol.	0.018	12:02
07 Room Air	0.000	12:02


08 Sim Temp = 34.0°C

Simul Ser No = 0000000

Std Sol No = 0000

County = 08

Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Low AC 0.021 AC

Form 106-I8000

*80-004954
25/31 FM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
03/10/2010 12:03

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:03
02 Std. Sol.	0.290	12:04
03 Room Air	0.000	12:05
04 Std. Sol.	0.290	12:06
05 Room Air	0.000	12:06
06 Std. Sol.	0.290	12:07
07 Room Air	0.000	12:08

08 Sim Temp = 34.0°C

Simul Ser No = 0000000
Std Sol No = 0000
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Linearity 0.285

Form 106-I8000

80-004954
26/31 FM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
03/10/2010 12:08

Test	AC	Time
01 Room Air	0.000	12:09
02 *Subject Test	INT*	12:09
03 Room Air	0.000	12:10

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 11/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Simul Ser No = 0000000
Std Sol No = 0000
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Alcohol 0.100AC + Acetone 0.5%

Form 106-I8000

80-004954
27/31 1912


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
03/10/2010 12:10

Test	AC	Time
01 Room Air	RFI*	12:11
02 Room Air	0.000	12:11

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 11/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 709352
Cyl No = 1
Expiration Date = 12/23/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

RFI

Form 106-I8000

80-004954
2/3/11

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
03/12/2010 10:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:29
02 Std. Gas	0.081	10:30
03 Room Air	0.000	10:30
04 Std. Gas	0.080	10:31
05 Room Air	0.000	10:31
06 Std. Gas	0.081	10:32
07 Room Air	0.000	10:32

Lot No = 709352
Cyl No = 1
Exp Date = 12/23/2011
County = 08 Oper No. = 107501

Margy Pearson

Operator Signature
MARGY PEARSON

Remarks:

Cal check ACA

Form 106-I8000

80-004954
29/31 YAL


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
03/12/2010 10:32

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:33
02 Std. Gas	0.081	10:33
03 Room Air	0.000	10:34
04 Std. Gas	0.081	10:34
05 Room Air	0.000	10:35
06 Std. Gas	0.080	10:35
07 Room Air	0.000	10:36

Lot No = 709352
Cyl No = 1
Exp Date = 12/23/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Cal Check ACA

Form 106-I8000

80-004954
3/31 YH


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
03/12/2010 10:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:36
02 Std. Gas	0.081	10:37
03 Room Air	0.000	10:37
04 Std. Gas	0.081	10:38
05 Room Air	0.000	10:38
06 Std. Gas	0.081	10:39
07 Room Air	0.000	10:39

Lot No = 709352
Cyl No = 1
Exp Date = 12/23/2011
County = 08 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Cal Check ACA

Form 106-I8000

80-004954
31/31 *SPR*