



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Shannon Liebold		
Location:	Mercer County Sheriff's Dept.		
Intoxilyzer® Serial Number:	<del>80-003061</del> 80-003060 DK		
Gas Lot Number:	659357	Gas Cylinder Number:	43
		Expiration Date:	9-3-11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)			
Test 1: 0. <u>081</u> %			
Test 2: 0. <u>081</u> %			
Test 3: 0. <u>081</u> %			
(ATTACH TEST RECORD)			

Shannon Liebold  
 Chemical Test Operator Signature

3-12-11  
 Date

Deb Kashner  
 Reviewed By

22 Mar 2011  
 Date

80-003060  
 P. 100 of 101  
 ysm

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = STAT      8164.13.00 06/09  
03/12/2011      03:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	03:09
02 Std. Gas	0.081	03:09
03 Room Air	0.000	03:10
04 Std. Gas	0.081	03:10
05 Room Air	0.000	03:10
06 Std. Gas	0.081	03:11
07 Room Air	0.000	03:11

Lot No = 659357

Cyl No = 43

Exp Date = 09/03/2011

County = 29

Oper No. = 020904



Operator Signature  
SHANNON P. LIEBOLD

Remarks:

Form 106-I8000

80-003060  
P. 102 of 101  
YAM