

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Lt. Tracy L. Meidinger # 258

Location: UND-PD

Intoxilyzer Serial No.: 80-004206

Lot Number: 659358

Cylinder Number: 31

Expiration Date: 09-03-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 083 %

Test 2: 0. 083 %

Test 3: 0. 082 %

(ATTACH TEST RECORD)

Lt. Tracy L. Meidinger # 258  
Chemical Test Operator Signature

03-14-2010  
Date

Deb Kashner  
Reviewed By

23 Mar 2010  
Date

80-004206  
P. 1 of 2

7/12

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004206  
Location = UNDP      8164.13.00 06/09  
03/14/2010      21:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:15
02 Std. Gas	0.083	21:15
03 Room Air	0.000	21:16
04 Std. Gas	0.083	21:16
05 Room Air	0.000	21:17
06 Std. Gas	0.082	21:17
07 Room Air	0.000	21:17

Lot No = 659358  
Cyl No = 31  
Exp Date = 09/03/2011  
County = 18      Oper No. = 119409

*Tracy L. Meidinger #258*  
Operator Signature  
TRACY L. MEIDINGER

Remarks:

Form 106-I8000

ACA  
03-14-10

80-004206  
P. 2 of 2  
7/2/12