

BREATH ALCOHOL INSTRUMENT CHAIN OF CUSTODY WORKSHEET

Instrument: 8000 Serial Number: 80-003061

Transported Out of the Crime Laboratory:

Agency Being Assigned Instrument: _____

Name/Agency of Person Transporting Instrument:

_____ (Print Name) _____ (Print Agency)

Signature of Person Receiving Instrument: _____

Date Receiving Instrument: ____/____/____ From: _____

Or Shipped Via:

USPS ____ UPS ____ FedEx ____ Other _____

Date Sending Instrument: ____/____/____ Air Bill No. _____

Complete Air Bill (Identify Instrument/Serial Number) and attach a copy to this worksheet.

Transported to the Crime Laboratory:

Agency Returning Instrument: CM1

Name/Agency of Person Returning Instrument:

_____ (Print Name) _____ (Print Agency)

Signature of Person Returning Instrument: _____

Date Returning Instrument: 16 / May / 2011 Received By: Deb Kashur

Or Received Via:

USPS ____ UPS ____ FedEx Other _____

Date Receiving Instrument: ____/____/____ Air Bill No. _____

Attach shipping paperwork (if any) to this worksheet.

Instrument Transfer Information Entered Into the Equipment Database:

Date of Action 16 / May / 11 By: Deb Kashur

80-003061
P. 116 of 118
Ym

Certificate of Calibration

Intoxilyzer® 8000

*This is to certify that the calibration of **Intoxilyzer**® serial number 80-003061, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.*

Date March 7 2011

Signed Knute Robinson
Technician

CMI INC.

80-003061
P. 1184/118
FM

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650519 CMI 7/9/09