



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Patrick Hudson

Location: Ward County Jail

Intoxilyzer Serial No.: 80-003058

Lot Number: 659358

Cylinder Number: 32

Expiration Date: 9/3/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 081 %

Test 2: 0. 081 %

Test 3: 0. 081 %

(ATTACH TEST RECORD)

Patrick Hudson  
 Chemical Test Operator Signature  
 Deb Kashner

3/17/2010  
 Date Cylinder Changed  
 23 Mar 2010

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-003058  
 P. 1 of 2  
 rm

80/81

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003058  
Location = WARD      8164.13.00 06/09  
03/17/2010      21:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:57
02 Std. Gas	0.081	21:58
03 Room Air	0.000	21:58
04 Std. Gas	0.081	21:58
05 Room Air	0.000	21:59
06 Std. Gas	0.081	21:59
07 Room Air	0.000	22:00

Lot No = 659358

Cyl No = 32

Exp Date = 09/03/2011

County = 51

Oper No. = 049509



Operator Signature  
PATRICK C. HUDSON

Remarks:

Form 106-I8000

80-003058  
P. J. J.  
JH  
8/1/11