



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Tyler A Wolfe</u>		
Location: <u>HARV</u>		
Intoxilyzer® Serial Number: <u>80-004935</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>45</u>	Expiration Date: <u>12/23/2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <u>0.077</u> % Test 2: <u>0.077</u> % Test 3: <u>0.077</u> %		
(ATTACH TEST RECORD)		

Tyler Wolfe
 Chemical Test Operator Signature

3/18/10 ACArunon 3/19/10 DK
 Date

Deb Kashner
 Reviewed By

29 Mar 2010
 Date

80-004935
 33/34 KM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = HARV 8164.13.00 06/09
03/19/2010 10:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:55
02 Std. Gas	0.077	10:56
03 Room Air	0.000	10:56
04 Std. Gas	0.077	10:56
05 Room Air	0.000	10:57
06 Std. Gas	0.077	10:57
07 Room Air	0.000	10:58

Lot No = 709352
Cyl No = 45
Exp Date = 12/23/2011
County = 52 Oper No. = 060901



Operator Signature
TYLER A. WOLFE

Remarks:

ACA Form 105-G

Form 106-I8000

80-004935
34/34 FK