



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (1-2010)

Serial Number: 80-004944 Instrument Location: CARR

Reason for Install/Repair:
 Install After Lab Repair/Inspection
 Other (Specify) _____

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® 8000.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: _____

Donita Row
 Field Inspector Signature

03/18/10
 Date

Deb Kashner
 Reviewed By

23 Mar 2010
 Date

80-004944
 34/37 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = CARR 8164.13.00 06/09
03/18/2010 16:45

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 709352
Standard Cyl #? 12
Standard Expiration? 12/23/2011
Oper No? 100754

Flow Cal. Date: 01/18/2010
Slope 693
Intercept -679634

IR Calibration Date: 01/18/2010

	3um	9um
0th Coef(*100):	-29206	-25219
1st Coef(*100):	257889	131140
2nd Coef(*100):	2394	1225
H2O adj(mg/l*10k):	734	452

***** Printer Test End *****



Operator Signature
DANIELLE M. ROSEWAREN

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = CARR 8164.13.00 06/09
03/18/2010 16:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:46
02 Std. Gas	0.080	16:47
03 Room Air	0.000	16:47
04 Std. Gas	0.081	16:48
05 Room Air	0.000	16:48
06 Std. Gas	0.079	16:48
07 Room Air	0.000	16:49

Lot No = 709352
Cyl No = 12
Exp Date = 12/23/2011
County = 16 Oper No. = 100754

Danielle Roseware

Operator Signature
DANIELLE M. ROSEWAREN

Remarks:

Form 106-I8000

80-004944
36/37 PM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = CARR 8164.13.00 06/09
03/18/2010 16:52

Test	AC	Time
01 Room Air	RFI*	16:53
02 Room Air	0.000	16:53

*Invalid Test
Inhibited - RFI

Sub Name = TUFTE, MICHAEL W
Sub DOB = 04/10/1953
Sub Sex = Male Weight = 240
Test = OTH Cit = 001
Dr. Lic. = ND/TUF539332
Lot No = 709352
Cyl No = 12
Expiration Date = 12/23/2011
County = 16 Oper No. = 100754



Operator Signature
DANIELLE M. ROSEWAREN

Remarks:

Form 106-I8000

80-004944
3/37 AM