



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Steven R. Fischer</u>		
Location: <u>Ward County S.O. 1 Jail, Minot</u>		
Intoxilyzer® Serial Number: <u>80-003066</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>32</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>081</u> % Test 2: 0. <u>081</u> % Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

Steven R. Fischer

Chemical Test Operator Signature

Deb Kashner

Reviewed By

3-18-11

Date

22 Mar ²¹ 2011

Date

80-003066
 p. 127 of 128
 JAM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = WARD 8164.13.00 06/09
03/18/2011 19:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:19
02 Std. Gas	0.081	19:19
03 Room Air	0.000	19:20
04 Std. Gas	0.081	19:20
05 Room Air	0.000	19:20
06 Std. Gas	0.081	19:21
07 Room Air	0.000	19:21

Lot No = 659358
Cyl No = 32
Exp Date = 09/03/2011
County = 51 Oper No. = 049904



Operator Signature
STEVEN R. FISCHER

Remarks:

Form 106-I8000

80-003066
P.128 of 128
FM