



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59281 (1-2010)

| | |
|---|----------------------------------|
| Serial Number: <u>80-004203</u> | Instrument Location: <u>COOP</u> |
| Reason for Install/Repair: | |
| <input checked="" type="checkbox"/> Install After Lab Repair/Inspection <input type="checkbox"/> Other (Specify) _____ | |

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® 8000.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: _____

Daniel Rom
Field Inspector Signature

03/18/11
Date

Deb Kashur
Reviewed By

22 Mar 2011
Date

80-004203
p. 97 of 100
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004203
Location = COOP 8164.13.00 06/09
03/18/2011 19:19

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05911080A1
Standard Cyl #? 5
Standard Expiration? 02/01/2013
Oper No? 100754

Flow Cal. Date: 01/28/2010
Slope 719
Intercept -711903

IR Calibration Date: 01/28/2010

| | 3um | 9um |
|--------------------|--------|--------|
| 0th Coef(*100): | -17469 | -28710 |
| 1st Coef(*100): | 264216 | 135022 |
| 2nd Coef(*100): | 2778 | 1081 |
| H2O adj(mg/l*10k): | 430 | 605 |

***** Printer Test End *****



Operator Signature
DANIELLE M. ROSEWAREN

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004203
Location = COOP 8164.13.00 06/09
03/18/2011 19:50

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 19:50 |
| 02 Std. Gas | 0.079 | 19:51 |
| 03 Room Air | 0.000 | 19:51 |
| 04 Std. Gas | 0.080 | 19:52 |
| 05 Room Air | 0.000 | 19:52 |
| 06 Std. Gas | 0.079 | 19:53 |
| 07 Room Air | 0.000 | 19:53 |

Lot No = 05911080A1
Cyl No = 5
Exp Date = 02/01/2013
County = 20 Oper No. = 100754



Operator Signature
DANIELLE M. ROSEWAREN

Remarks:

Form 106-I8000

80-004203
P. 99 of 100
VFM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004203
Location = COOP 8164.13.00 06/09
03/18/2011 19:41

| Test | AC | Time |
|---------------|------|-------|
| 01 Diagnostic | OK | 19:43 |
| 02 Room Air | RFI* | 19:43 |
| 03 Room Air | RFI* | 19:43 |

*Invalid Test
Inhibited - RFI

Sub Name = TEST, NA NA
Sub DOB = 10/12/1976
Sub Sex = Female Weight = 150
Test = DUI Cit = 507689
Dr. Lic. = ND/NA
Lot No = 05911080A1
Cyl No = 5
Expiration Date = 02/01/2013
County = 20 Oper No. = 100754

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DANIELLE M. ROSEWAREN

Remarks:

Form 106-I8000

80-004203
p. 100 of 100
km