



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Danielle M. Rosewarem</u>		
Location: <u>COOP</u>		
Intoxilyzer® Serial Number: <u>80-004203</u>		
Gas Lot Number: <u>05711080A1</u>	Gas Cylinder Number: <u>5</u>	Expiration Date: <u>02/01/2013</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.080</u> %		
Test 2: <u>0.080</u> %		
Test 3: <u>0.079</u> %		
<b>(ATTACH TEST RECORD)</b>		

Danielle Rosewarem  
Chemical Test Operator Signature

03/18/11  
Date

Deb Kashur  
Reviewed By

22 Mar 2011  
Date


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004203  
Location = COOP      8164.13.00 06/09  
03/18/2011      19:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:47
02 Std. Gas	0.080	19:47
03 Room Air	0.000	19:48
04 Std. Gas	0.080	19:48
05 Room Air	0.000	19:48
06 Std. Gas	0.079	19:49
07 Room Air	0.000	19:49

Lot No = 05911080A1  
Cyl No = 5  
Exp Date = 02/01/2013  
County = 20      Oper No. = 100754

  
\_\_\_\_\_  
Operator Signature  
DANIELLE M. ROSEWAREN

Remarks:

Form 106-I8000

80-004203  
p. 103 of 102  
FM