



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Duane L. Marmor				
Location:	STAN				
Intoxilyzer® Serial Number:	80-004943				
Gas Lot Number:	709352	Gas Cylinder Number:	34	Expiration Date:	12/23/11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0. <u>079</u> %					
Test 2: 0. <u>079</u> %					
Test 3: 0. <u>079</u> %					
(ATTACH TEST RECORD)					

Duane L. Marmor
Chemical Test Operator Signature

03/19/11
Date

Deb Kashner
Reviewed By

22 Mar 2011
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
03/19/2011 19:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:08
02 Std. Gas	0.079	19:08
03 Room Air	0.000	19:09
04 Std. Gas	0.079	19:09
05 Room Air	0.000	19:10
06 Std. Gas	0.079	19:10
07 Room Air	0.000	19:11

Lot No = 709352
Cyl No = 34
Exp Date = 12/23/2011
County = 31 Oper No. = 100019


Operator Signature
DUANE MARMON

Remarks:

Form 106-I8000

80-004943
P-70 of 70
FM