



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Todd A. Lerol				
Location:	GRAF				
Intoxilyzer® Serial Number:	80-003072				
Gas Lot Number:	659358	Gas Cylinder Number:	16	Expiration Date:	09/03/2011
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.082%					
Test 2: 0.082%					
Test 3: 0.082%					
(ATTACH TEST RECORD)					

Chemical Test Operator Signature

Reviewed By

03/20/2010 DK
~~03/29/2010~~

Date

29 Mar 2010

Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = GRAF 8164.13.00 06/09
03/20/2010 16:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:39
02 Std. Gas	0.082	16:40
03 Room Air	0.000	16:40
04 Std. Gas	0.082	16:40
05 Room Air	0.000	16:41
06 Std. Gas	0.082	16:41
07 Room Air	0.000	16:42

Lot No = 659358
Cyl No = 16
Exp Date = 09/03/2011
County = 50 Oper No. = 019709



Operator Signature
TODD LEROL

Remarks:

Form 106-I8000

80-003072
P.2 of 2 VCL
74/75