



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Danielle Rosevaren</u>		
Location: <u>CARR</u>		
Intoxilyzer® Serial Number: <u>80-004944</u>		
Gas Lot Number: <u>7009352</u>	Gas Cylinder Number: <u>12</u>	Expiration Date: <u>12/23/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>080</u> % Test 2: 0. <u>079</u> % Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

Danielle Rosevaren  
Chemical Test Operator Signature

03/20/11  
Date

Deb Kasher  
Reviewed By

25 Mar 2011  
Date

80-004944  
P. 61462  
sam

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004944  
Location = CARR                8164.13.00 06/09  
03/20/2011                        21:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:20
02 Std. Gas	0.080	21:20
03 Room Air	0.000	21:21
04 Std. Gas	0.079	21:21
05 Room Air	0.000	21:22
06 Std. Gas	0.079	21:22
07 Room Air	0.000	21:23

Lot No = 709352  
Cyl No = 12  
Exp Date = 12/23/2011  
County = 16                      Oper No. = 100754

  
\_\_\_\_\_  
Operator Signature  
DANIELLE ROSEWAREN

Remarks:

Form 106-I8000

80-004944  
P. 62 of 62  
RM