



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Chad Harmon</u>		
Location: <u>FYPD</u>		
Intoxilyzer® Serial Number: <u>80-004954</u>		
Gas Lot Number: <u>15410080A1</u>	Gas Cylinder Number: <u>34</u>	Expiration Date: <u>07-01-12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>079</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

Chad Harmon
Chemical Test Operator Signature

03-20-11
Date

Deb Kashur
Reviewed By

14 Apr 2011
Date

80-004954
P. 67 of 68
JAM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = FYPD 8164.13.00 06/09
03/20/2011 11:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:07
02 Std. Gas	0.079	11:07
03 Room Air	0.000	11:07
04 Std. Gas	0.079	11:08
05 Room Air	0.000	11:08
06 Std. Gas	0.079	11:09
07 Room Air	0.000	11:09

Lot No = 15410080A1
Cyl No = 34
Exp Date = 07/01/2012
County = 61 Oper No. = 010606



Operator Signature
CHAD L. HARMON

Remarks:

Form 106-I8000

80-004954
P. 28 of 68
vrm