

INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006504

A. GENERAL Setup and Checks:

- Power On
 5 min. countdown during initial warm-up
 1 min. countdown to begin test after standby mode
- Display Intensity (Ctrl + ↑ or ↓)
3. Password Security:
 Level 1 (Operator) Level 2 (Field Inspector) Level 3 (Lab Personnel)
4. Printer Setup (Level 3,S,P)
5. Forms Load (CMI, Inc., CoBrA Software)

B. LEVEL 1 Setup and Checks:

- Internal Printer Test and Paper Feed (Level 1,P). Attach record.
- External Printer Test (Level 1,P). Attach record.
3. Barcode Reader (Level 1,B):
 Operator ID ND License R key: Refusal. Attach record.

C. LEVEL 2 Setup and Checks:

- Continuous Air Blank (Level 2,A)
- Diagnostic (Level 2,D). Attach record.
- Prelim Data Entry (Level 2,E)
4. Ser Number/Version Info (Level 2,V):
Inst Ser#: 80-006504 DSP Ver: 1362.14
Core Ver: 8164.13 CPLD Ver: 7

D. LEVEL 3 Setup and Checks:

- DVM Monitor (Level 3,D,D): 3: 12440 s: 1 C: 47.00
9: 13052 s: 0 B: 45 0
(RFI indicator with keyed radio): 1
2. Function (F) Keys: : F5 – Tone F9 - Fan
 F6 – Green LED F10 – Pump
 F7 – Red LED F11 – Solenoid
3. Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 955 CF: 1.06
4. Flow Monitor (Level 3,D,F): Flow Sensor, R: 199 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 50.0 I: 0.0
B: 45.0 V: 0
6. Tank Monitor (Level 3,D,G): P (display): 960 psi P (gauge): 950 psi
7. General Setup (Level 3,S,G)
8. Configure Start Test (Level 3,S,T)
9. Network (Level 3,S,N)

E. TESTS (attach all test records):

Configure wet bath simulator standards before running tests (Level 1,S):

1. Zero AC (Level 1,C):
Sim ser #: DR3375 Lot #: Milli-Q Exp date: N/A
Known value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):
Sim ser #: DR3374 Lot #: 14240 Exp date: 10/20/16
Known value: 0.020
Results: 0.020 AC 0.020 AC 0.021 AC
3. High AC (Level 1,C):
Sim ser #: DR7111 Lot #: 201408C Exp date: 08/12/16
Known value: 0.300
Results: 0.303 AC 0.302 AC 0.302 AC

Configure dry gas standard before running tests (Level 1,S):

4. Ethanol + 0.05% Acetone (Level 1,B):
Sim ser #: DR5144 Lot #: ICS4 Exp date: 09/09/17
Known value: 0.10 AC + 0.05% acetone
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):
Sim ser #: DR3455 Lot #: Acet 1 Exp date: 09/29/17
Known value: 0.05% acetone
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):
Sim ser #: DR5189 Lot #: Met 2 Exp date: 09/29/17
Known value: 0.10% Methanol
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):
Sim ser #: DR3454 Lot #: IPA 2 Exp date: 09/29/17
Known value: 0.10% isopropanol
 Display reads "Interferent Detect"

- 8. Ambient Fail (Level 1,B): Display and test record read correct.
- 9. Invalid Sample (Level 1,B): Display and test record read correct.
- 10. Sequence Aborted (Level 1,B): Display and test record read correct.
- 11. ABA (Level 1,B): Test record reads correct.
- 12. Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13. RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14. Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15. Deficient Samples (CMS mode): Display and test record read correct.
- 16. Improper Sample (CMS mode): Display and test record read correct.
- 17. Difference Too Great (CMS mode): Display and test record read correct.
- 18. CMS (CMS mode): Test record reads correct.
- 19. Stability Test (Level 3,T,S):

Ave: 0.0802 Std Dev: 0.0005 Rel Std Dev: 0.10270

Maintenance/Repairs/Notes: Initial inspection - realigned internal printer

Instrument is acceptable to be used in the field.

Deb Shanaver
Field Inspector's Signature

24 Mar 2016
Date

Charles E. Eda
Reviewed by

24 MAR 16
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 07:16

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

24 characters per line
abcdefghijklmnopqrstuwx
ABCDEFGHIJKLMNQRSTUWX

32 characters per line
abcdefghijklmnopqrstuwx123456
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuwx1234567890abcdef
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup
Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On MemFull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 80
Standard Cyl #? 1234567
Standard Expiration? 11/11/2020
Oper No? 777777

Flow Cal. Date: 08/20/2015
Slope 681
Intercept -673151

IR Calibration Date: 08/20/2015
3um 9um

0th Coef(*100): -14464 -14817
1st Coef(*100): 255017 131849
2nd Coef(*100): 5533 1946
H2O adj(mg/l*10K): 700 431

***** Test End *****

Deb Shanauer
Operator Signature
DEB SHANAUER

Remarks:

Print test

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 08:10

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=-|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=-|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 80
Standard Cyl #? 1234567
Standard Expiration? 11/11/2020
Oper No? 777777

Flow Cal. Date: 08/20/2015
Slope 681
Intercept -673151

IR Calibration Date: 08/20/2015

	3um	9um
0th Coef(*100):	-14464	-14817
1st Coef(*100):	255017	131849
2nd Coef(*100):	5533	1946
H2O adj(mg/l*10k):	700	431

***** Printer Test End *****


Operator Signature
DEB SHANAVER

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 08:11

Test	AC	Time
01 Room Air	0.000	08:12
02 *Subject Test	REF*	08:12
03 Room Air	0.000	08:12


*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 80
Cyl No = 1234567
Expiration Date = 11/11/2020
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:



Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 08:21

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Diagnostic

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 08:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:27
02 Std. Sol.	0.000	08:27
03 Room Air	0.000	08:28
04 Std. Sol.	0.000	08:29
05 Room Air	0.000	08:29
06 Std. Sol.	0.000	08:30
07 Room Air	0.000	08:30

08 Sim Temp = 34.0°C

Simul Ser No = DR3375
Std Sol No = MILLIQ
County = 08 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

zero alcohol 0.000 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 08:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:36
02 Std. Sol.	0.020	08:38
03 Room Air	0.000	08:38
04 Std. Sol.	0.020	08:39
05 Room Air	0.000	08:40
06 Std. Sol.	0.021	08:40
07 Room Air	0.000	08:41

08 Sim Temp = 34.0°C

Simul Ser No = DR3374

Std Sol No = 14240

County = 08

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

low alcohol 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 08:50

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:51
02 Std. Sol.	0.303	08:52
03 Room Air	0.000	08:52
04 Std. Sol.	0.302	08:53
05 Room Air	0.000	08:53
06 Std. Sol.	0.302	08:54
07 Room Air	0.000	08:55


08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 201408C

County = 08

Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

High alcohol 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 09:07

Test	AC	Time
01 Diagnostic	OK	09:08
02 Room Air	0.000	09:09
03 *Subject Test	INT*	09:09
04 Room Air	0.000	09:10

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.10AC + 0.05% acetone

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 09:11

Test	AC	Time
01 Diagnostic	OK	09:11
02 Room Air	0.000	09:12
03 *Subject Test	INT*	09:12
04 Room Air	0.000	09:13

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.05% acetone

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 09:19

Test	AC	Time
01 Diagnostic	OK	09:20
02 Room Air	0.000	09:20
03 *Subject Test	INT*	09:20
04 Room Air	0.000	09:21

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.10% methanol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 09:22

Test	AC	Time
01 Diagnostic	OK	09:23
02 Room Air	0.000	09:24
03 *Subject Test	INT*	09:24
04 Room Air	0.000	09:25

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Interferent detected

0.10% isopropanol

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 10:36

Test	AC	Time
01 Room Air	AMB*	10:36
02 Room Air	PUR**	10:37

*Invalid Test
 Check Ambient Conditions
**Purge Fail

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Ambient fail (plus Purge fail)

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 10:41

Test	AC	Time
01 Room Air	0.000	10:42
02 Invalid Sample	X.XXX	10:42
03 Room Air	0.000	10:43

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Invalid sample - mouth alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 10:43

Test	AC	Time
01 Room Air	0.000	10:44
02 *Subject Test	ABT*	10:44
03 Room Air	0.000	10:45

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Sequence aborted

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 10:39

Test	AC	Time
01 Room Air	0.000	10:40
02 Subject Test 1	0.195	10:40
03 Room Air	0.000	10:41
04 Reported AC	0.195	10:40

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

ABA

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 08:57

Test	AC	Time
01 Diagnostic	OK	08:58
02 Room Air	0.000	08:58
03 Subject Test 1	0.000	08:58
04 Room Air	0.000	09:01
05 Std. Gas	0.101*	09:02
06 Room Air	0.000	09:02

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

Calibration check out of tolerance

Intoxilyzer Test Record and Checklist
 NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer	Alcohol Analyzer
North Dakota Model 8000	SN 80-006504
Location = TOXL	8164.13.00 06/09
03/24/2016	10:46

Test	AC	Time
01 Diagnostic	OK	10:47
02 Room Air	RFI*	10:47
03 Room Air	0.000	10:47

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
 Sub DOB = 02/01/1992
 Sub Sex = Female
 Weight = 150
 Test = DUI
 Cit = NA
 Dr. Lic. = ND/DIS921456
 Lot No = 19615080A5
 Cyl No = 26
 Expiration Date = 09/05/2017
 County = 08
 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
 DEB SHANAVER

Remarks:

RFI

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 13:25

Test	AC	Time
01 Diagnostic	OK	13:26
02 Room Air	0.000	13:26
03 *Subject Test	0.000*	13:29
04 Room Air	0.000	13:31
05 Std. Gas	0.080	13:33
06 Room Air	0.000	13:34
07 *Subject Test	0.000*	13:37
08 Room Air	0.000	13:38

No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:

Deficient samples

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 13:48

Test	AC	Time
01 Diagnostic	OK	13:49
02 Room Air	0.000	13:49
03 Subject Test 1	0.000	13:50
04 Room Air	0.000	13:52
05 Std. Gas	0.078	13:53
06 Room Air	IPS*	13:53
07 Room Air	0.000	13:54

*Improper Sample

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.


Operator Signature
DEB SHANAVER

Remarks:

Improper sample

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 13:55

Test	AC	Time
01 Diagnostic	OK	13:55
02 Room Air	0.000	13:56
03 Subject Test 1	0.039*	13:56
04 Room Air	0.000	13:59
05 Std. Gas	0.080	14:00
06 Room Air	0.000	14:01
07 Subject Test 2	0.000*	14:01
08 Room Air	0.000	14:02

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Difference too great

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 14:03

Test	AC	Time
01 Diagnostic	OK	14:04
02 Room Air	0.000	14:04
03 Subject Test 1	0.000	14:05
04 Room Air	0.000	14:07
05 Std. Gas	0.079	14:08
06 Room Air	0.000	14:09
07 Subject Test 2	0.000	14:09
08 Room Air	0.000	14:10
09 Reported AC	0.000	14:05

Difference OK
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 19615080A5
Cyl No = 26
Expiration Date = 09/05/2017
County = 08 Oper No. = 777777

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:

CMS

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006504
Location = TOXL 8164.13.00 06/09
03/24/2016 14:12

Test	AC	Time
01 Room Air	0.000	14:12
02 Std. Gas	0.080	14:12
03 Room Air	0.000	14:13
04 Std. Gas	0.081	14:13
05 Room Air	0.000	14:14
06 Std. Gas	0.080	14:14
07 Room Air	0.000	14:15
08 Std. Gas	0.080	14:15
09 Room Air	0.000	14:16
10 Std. Gas	0.079	14:16
11 Room Air	0.000	14:16
12 Std. Gas	0.080	14:17
13 Room Air	0.000	14:17
14 Std. Gas	0.080	14:18
15 Room Air	0.000	14:18
16 Std. Gas	0.080	14:19
17 Room Air	0.000	14:19
18 Std. Gas	0.080	14:19
19 Room Air	0.000	14:20
20 Std. Gas	0.080	14:20
21 Room Air	0.000	14:21
22 Std. Gas	0.081	14:21
23 Room Air	0.000	14:22
24 Std. Gas	0.081	14:22
25 Room Air	0.000	14:23
26 Std. Gas	0.080	14:23
27 Room Air	0.000	14:24
28 Std. Gas	0.081	14:24
29 Room Air	0.000	14:24
30 Std. Gas	0.080	14:25
31 Room Air	0.000	14:25
32 Std. Gas	0.080	14:26
33 Room Air	0.000	14:26
34 Std. Gas	0.080	14:27
35 Room Air	0.000	14:27
36 Std. Gas	0.080	14:27
37 Room Air	0.000	14:28
38 Std. Gas	0.080	14:28
39 Room Air	0.000	14:29
40 Std. Gas	0.080	14:29
41 Room Air	0.000	14:30
42 Cal Check Stats:		
43 Average	0.0802	
44 Std Dev	0.0005	
45 Rel Std Dev	0.62%	

*Stability Test
Deb Shanaver
p. 1 of 2*

Operator Signature
DEB SHANAVER

Remarks:

Lot #: 19615080A5

cyl #: 26

Exp: 09/05/2017

Form 106-I8000

Stability test
Deb Shanaver
P. 2 of 2