



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION

Toxicology Section/Breath Alcohol Program

SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Phil Wolf</u>		
Location: <u>Cavalier, MD</u>		
Intoxilyzer® Serial Number: <u>80-002669</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>9</u>	Expiration Date: <u>9/3/11</u>
<input type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>077</u> %		
Test 2: 0. <u>077</u> %		
Test 3: 0. <u>077</u> %		
(ATTACH TEST RECORD)		

P. Wolf
Chemical Test Operator Signature

3/25/11
Date

Deb Kashner
Reviewed By

14 Apr 11
Date

80-002669
P. 106 of 107
FM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002669
Location = CAVA 8164.13.00 06/09
03/25/2011 03:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	03:21
02 Std. Gas	0.077	03:21
03 Room Air	0.000	03:22
04 Std. Gas	0.077	03:22
05 Room Air	0.000	03:23
06 Std. Gas	0.077	03:23
07 Room Air	0.000	03:23

Lot No = 659357
Cyl No = 9
Exp Date = 09/03/2011
County = 34 Oper No. = 120547



Operator Signature
PHIL WOLF

Remarks:

Form 106-I8000

80-002669
p. 107 of 107
FM