



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>RAY Weber</u>		
Location: <u>TRAIL CO. S.O.</u>		
Intoxilyzer® Serial Number: <u>80-004949</u>		
Gas Lot Number: <u>154/0080A1</u>	Gas Cylinder Number: <u>#15</u>	Expiration Date: <u>07/01/12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>080</u> % Test 2: 0. <u>079</u> % Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Ray Weber
 Chemical Test Operator Signature

3/25/11
 Date

Deb Kashner
 Reviewed By

17 Apr 11
 Date

80-004949
 P. 98-199
 VFL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004949
Location = HILL 8164.13.00 06/09
03/25/2011 22:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:22
02 Std. Gas	0.080	22:22
03 Room Air	0.000	22:23
04 Std. Gas	0.079	22:23
05 Room Air	0.000	22:24
06 Std. Gas	0.080	22:24
07 Room Air	0.000	22:24

Lot No = 15410080A1
Cyl No = 15
Exp Date = 07/01/2012
County = 49 Oper No. = 058405



Operator Signature
RAY G. WEBER

Remarks:

Form 106-I8000

80-004949
P. 99 of 99
RM