



SERVICE WORK ORDER

DATE: 3-26-15

RCVD VIA: FEDEX

316 E. 9th St., Owensboro KY 42303
 Phone 866-835-0690
 Fax: 270-685-6268

336199

BILL TO:

ATTN:

CITY STATE ZIP

CUSTOMER NO. 585013

PHONE

SHIP TO: ND Crime Lab

ATTN: Deb Kasher

2641 E. Main Ave.

CITY: Bismarck STATE: ND ZIP: 58501

PHONE: 701-328-6159

FAX: 701-328-6485

MODEL # 8000 SERIAL # 80-005953 BILL CODE EW

EXTRA PARTS RCVD: 1st, probly Battery

DESCRIPTION OF PROBLEM: EEPROM Fail. System fault #18

Infrared / Fuel Cell
 PRELIMINARY TEST INFORMATION

System Error 18

BATTERY-EXT 12.46V, INT 3.04V

VERIFIED PROBLEM? Yes

WORK PERFORMED:
 Cleared system error
 Performed system checks - OK
 No parts used

PARTS USED		
QTY	PART NO.	DESCRIPTION

Miscellaneous
 PRELIMINARY TEST INFORMATION

WORK PERFORMED:

[Signature]

REPAIR GTW	CAL. ADJ. DR	CALIBRATOR DR	FINAL DR
1/2 HOURS 4	HOURS 1	HOURS 1	HOURS 1
DATE Apr 27, 2015	DATE Apr 30, 2015	DATE May 1, 2014	DATE May 4, 2015

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS