

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-005953

Agency being assigned instrument: CM1

Date 24 Mar 15

Transferred from:

Transferred to:

Name (print) Deb Kashur

Name (print) CM1

Signature Deb Kashur

Signature _____

Agency Crime Lab

Agency _____

Reason for transfer/remarks:

EEPROM Fail
System Fault
ID#: 018
under warranty

Or Shipped Via:

USPS _____ UPS _____ FedEx Other _____

Date Sending Instrument: 24, Mar, 15 Air Bill No. _____



Service Evaluation Form

This form **MUST** be completed and enclosed with instrument to be serviced.
Failure to complete and return this form may cause delays in service.

(Note: please ship items in their original shipping container(s) or a similar protective box.)

1. Contact information: Customer Number 585013 (contact Customer Service)
 Name Deb Kashur Phone: (701) 328-6159
 Fax: (701) 328-6185 Email: dkashur@nd.gov

2. Bill to Address: under warranty Ship to Address: N.D. Crime Laboratory
2641 E. Main Ave
Bismarck, ND 58501-5044

3. Serial Number: 80-005953 Instrument Model: 8000

4. Detailed Description of Problem:
EEPROM Fail data invalid
System fault
ID#: 018 (power cord not attached/packed)
*****Hazardous Material Warning! – DO NOT return gas cylinders with instrument!*****

I Authorize Repairs Up To: All \$250 \$500 \$750 Other \$ _____
 Purchase Order Number (attach a copy of P.O. if applicable) _____

Authorized By: under warranty

Name (Please Print) _____ Title _____
 Signature _____ Date _____

Ship to:
 CMI, Inc.
 Attn: Service Dept.
 316 East Ninth Street
 Owensboro, KY 42303

No, please send estimate before repairs are made.
 Note: An estimate will be faxed before performing any repairs and may cause delays in service.
 An evaluation fee (\$79.00 or actual costs) will apply to estimates that are not repaired.