

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Steven R Fischer

Location: Rolette County Sheriff's Office, Rolla, ND

Intoxilyzer Serial No.: 80-004199

Lot Number: 659358

Cylinder Number: 11

Expiration Date: 09-03-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 081 %

Test 2: 0. 080 %

Test 3: 0. 082 %

(ATTACH TEST RECORD)

  
Chemical Test Operator Signature

3-29-2010  
Date

Deb Kashum  
Reviewed By

07 Apr 2010  
Date

80-004199  
7/73 km

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004199  
Location = ROLL      8164.13.00 06/09  
03/29/2010      01:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:48
02 Std. Gas	0.081	01:48
03 Room Air	0.000	01:49
04 Std. Gas	0.080	01:49
05 Room Air	0.000	01:50
06 Std. Gas	0.082	01:50
07 Room Air	0.000	01:51

Lot No = 659358

Cyl No = 11

Exp Date = 09/03/2011

County = 40

Oper No. = 049904



Operator Signature  
STEVEN R. FISCHER

Remarks:

Form 106-I8000

80-004199  
73/73 192