



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Darcy W Aberle</u>		
Location: <u>Williams County LEC</u>		
Intoxilyzer® Serial Number: <u>80-004184</u>		
Gas Lot Number: <u>15410080A1</u>	Gas Cylinder Number: <u>6</u>	Expiration Date: <u>7/01/12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.080</u> %		
Test 2: <u>0.080</u> %		
Test 3: <u>0.081</u> %		
(ATTACH TEST RECORD)		

Darcy W Aberle  
Chemical Test Operator Signature

03/29/11  
Date

Deb Kasher  
Reviewed By

14 Apr 11  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004184  
Location = WILL      8164.13.00 06/09  
03/29/2011      12:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:56
02 Std. Gas	0.080	12:56
03 Room Air	0.000	12:57
04 Std. Gas	0.080	12:57
05 Room Air	0.000	12:58
06 Std. Gas	0.081	12:58
07 Room Air	0.000	12:58

Lot No = 15410080A1  
Cyl No = 6  
Exp Date = 07/01/2012  
County = 53      Oper No. = 010101

*Darcy W Aberle*

Operator Signature  
DARCY W. ABERLE

Remarks:

Form 106-I8000

80-004184  
P. 107 of 107  
FM