

INTOXILYZER® 8000 INITIAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006501

A. GENERAL Setup and Checks:

- Power On
 - 5 min. countdown during initial warm-up
 - 1 min. countdown to begin test after standby mode
- Display Intensity (Ctrl + ↑ or ↓)
- Password Security:
 - Level 1 (Operator) Level 2 (Field Inspector) Level 3 (Lab Personnel)
- Printer Setup (Level 3,S,P)
- Forms Load (CMI, Inc., CoBrA Software)

B. LEVEL 1 Setup and Checks:

- Internal Printer Test and Paper Feed (Level 1,P). Attach record.
- External Printer Test (Level 1,P). Attach record.
- Barcode Reader (Level 1,B):
 - Operator ID ND License R key: Refusal. Attach record.

C. LEVEL 2 Setup and Checks:

- Continuous Air Blank (Level 2,A)
- Diagnostic (Level 2,D). Attach record.
- Prelim Data Entry (Level 2,E)
- Ser Number/Version Info (Level 2,V):
 - Inst Ser#: 80-006501 DSP Ver: 13762.14^{CE}
 - Core Ver: 8164.13 CPLD Ver: 7

D. LEVEL 3 Setup and Checks:

- DVM Monitor (Level 3,D,D): 3: 12556 s: 1 c: 47.00
9: 13286 s: 1 B: 45 0
(RFI indicator with keyed radio): 1
- Function (F) Keys: : F5 – Tone F9 - Fan
 F6 – Green LED F10 – Pump
 F7 – Red LED F11 – Solenoid
- Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 952 CF: 1.06
- Flow Monitor (Level 3,D,F): Flow Sensor, R: 234 F: 0.000 L/s

5. Temperature Monitor (Level 3,D,T): C: 47.0 A: 48.5 I: 0.0
B: 45.0 V: 0.0
6. Tank Monitor (Level 3,D,G): P (display): 1005 psi P (gauge): 21000 psi
7. General Setup (Level 3,S,G)
8. Configure Start Test (Level 3,S,T)
9. Network (Level 3,S,N)

E. TESTS (attach all test records):

Configure wet bath simulator standards before running tests (Level 1,S):

1. Zero AC (Level 1,C):
Sim ser #: MP3067 Lot #: MILLI Q H₂O Exp date: NA
Known value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC
2. Low AC (Level 1,C):
Sim ser #: MP3066 Lot #: 15160 Exp date: 7.14.17
Known value: 0.020 AC
Results: 0.016 AC 0.017 AC 0.016 AC
3. High AC (Level 1,C):
Sim ser #: MP3068 Lot #: 14102 Exp date: 4.23.16
Known value: 0.400 AC
Results: 0.408 AC 0.408 AC 0.409 AC

Configure dry gas standard before running tests (Level 1,S):

4. Ethanol + 0.05% Acetone (Level 1,B):
Sim ser #: DR5135 Lot #: ICS 4 Exp date: 9.9.17
Known value: 0.10 AC + 0.05% ACETONE
 Display reads "Interferent Detect"
5. Acetone (Level 1,B):
Sim ser #: DR5143 Lot #: ACE 2 Exp date: 9.29.17
Known value: 0.05% ACETONE
 Display reads "Interferent Detect"
6. Methanol (Level 1,B):
Sim ser #: DR3377 Lot #: MET 2 Exp date: 9.29.17
Known value: 0.10% METHANOL
 Display reads "Interferent Detect"
7. Isopropanol (Level 1,B):
Sim ser #: DR3453 Lot #: IPA 2 Exp date: 9.29.17
Known value: 0.10% ISOPROPANOL
 Display reads "Interferent Detect"

- 8. Ambient Fail (Level 1,B): Display and test record read correct.
- 9. Invalid Sample (Level 1,B): Display and test record read correct.
- 10. Sequence Aborted (Level 1,B): Display and test record read correct.
- 11. ABA (Level 1,B): Test record reads correct.
- 12. Calibration Check Out of Tolerance (CMS mode): Display and test record read correct.
- 13. RFI Detected (B,C, or CMS mode): Display and test record read correct.
- 14. Tank Pressure Below Minimum (CMS mode): Display reads correct.
- 15. Deficient Samples (CMS mode): Display and test record read correct.
- 16. Improper Sample (CMS mode): Display and test record read correct.
- 17. Difference Too Great (CMS mode): Display and test record read correct.
- 18. CMS (CMS mode): Test record reads correct.
- 19. Stability Test (Level 3,T,S):

Ave: 0.081 Std Dev: 0.0006 Rel Std Dev: 0.70%

Maintenance/Repairs/Notes: _____

Instrument is acceptable to be used in the field.

Charles E. Edm
Field Inspector's Signature

29 MAR 16
Date

Deb Shanaver
Reviewed by

29 Mar 16
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/28/2016 16:28

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

24 characters per line
abcdefghijklmnopqrstuvw
ABCDEFGHIJKLMNQRSTUW

32 characters per line
abcdefghijklmnopqrstuvwxy123456
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuvwxy1234567890
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEFGHI

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABHLRBH
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Menfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 1234567
Standard Cyl #? 13568
Standard Expiration? 11/11/2020
Oper No? 666666

Flow Cal. Date: 08/18/2015
Slope 670
Intercept -629984

IR Calibration Date: 08/18/2015
3um 9um

0th Coef(*100):	-18066	-17820
1st Coef(*100):	261817	138262
2nd Coef(*100):	4970	1663
H2O adj(mg/l*10K):	666	427

***** Test End *****



Operator Signature
CHARLES EDER

Remarks: **INTERNAL
PRINT TEST**

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/28/2016 16:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 29614080A2
Standard Cyl #? 6
Standard Expiration? 12/05/2016
Oper No? 666666

Flow Cal. Date: 08/18/2015
Slope 670
Intercept -629984

IR Calibration Date: 08/18/2015
 3um 9um

0th Coef(*100): -18086 -17820
1st Coef(*100): 261817 138262
2nd Coef(*100): 4970 1663
H2O adj(mg/l*10k): 666 427

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

EXTERNAL PRINT
TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/28/2016 16:42

Test	AC	Time
01 Room Air	0.000	16:43
02 *Subject Test	REF*	16:43
03 Room Air	0.000	16:43

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = REFUSAL
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

REFUSAL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/28/2016 16:44

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

DIAGNOSTIC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 10:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:25
02 Std. Sol.	0.000	10:26
03 Room Air	0.000	10:26
04 Std. Sol.	0.000	10:27
05 Room Air	0.000	10:27
06 Std. Sol.	0.000	10:28
07 Room Air	0.000	10:28

08 Sim Temp = 34.0°C

Simul Ser No = MP3067
Std Sol No = MILLIQ H2O
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

0.000 AC
ZERO AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 10:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:32
02 Std. Sol.	0.016	10:32
03 Room Air	0.000	10:33
04 Std. Sol.	0.017	10:34
05 Room Air	0.000	10:34
06 Std. Sol.	0.016	10:35
07 Room Air	0.000	10:36

08 Sim Temp = 34.0°C

Simul Ser No = MP3066
Std Sol No = 15160
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 10:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:45
02 Std. Sol.	0.408	10:46
03 Room Air	0.000	10:47
04 Std. Sol.	0.408	10:48
05 Room Air	0.000	10:48
06 Std. Sol.	0.409	10:49
07 Room Air	0.000	10:50

08 Sim Temp = 34.0°C

Simul Ser No = MP3068
Std Sol No = 14102
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.400 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 10:56

Test	AC	Time
01 Room Air	0.000	10:56
02 *Subject Test	INT*	10:57
03 Room Air	0.000	10:58

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: **INTERFERENT**
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 11:00

Test	AC	Time
01 Room Air	0.000	11:00
02 *Subject Test	INT*	11:01
03 Room Air	0.000	11:02

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INTERFERENT
0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 11:03

Test	AC	Time
01 Room Air	0.000	11:04
02 *Subject Test	INT*	11:04
03 Room Air	0.000	11:05

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT*
0.10% METHANOL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 11:06

Test	AC	Time
01 Room Air	0.000	11:07
02 *Subject Test	INT*	11:07
03 Room Air	0.000	11:08

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INTERFERENT
0.10% ISOPROPNOL

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 11:09

Test	AC	Time
01 Room Air	AMB*	11:10
02 Room Air	0.000	11:10

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = AMBIENT FAIL
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *AMBIENT
FAIL*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 11:11

Test	AC	Time
01 Room Air	0.000	11:12
02 Invalid Sample	X.XXX	11:12
03 Room Air	0.000	11:13

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INVALID SAMPLE
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

*INVALID
SAMPLE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 11:13

Test	AC	Time
01 Room Air	0.000	11:14
02 *Subject Test	ABT*	11:14
03 Room Air	0.000	11:15

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = SEQ ABORTED
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: SEQUENCE
ABORTED

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 11:16

Test	AC	Time
01 Room Air	0.000	11:17
02 Subject Test 1	0.010	11:17
03 Room Air	0.000	11:17
04 Reported AC	0.010	11:17

No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = ABA TEST
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ABA TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 12:29

Test	AC	Time
01 Diagnostic	OK	12:30
02 Room Air	0.000	12:30
03 Subject Test 1	0.000	12:31
04 Room Air	0.000	12:33
05 Std. Gas	0.087*	12:34
06 Room Air	0.000	12:35

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = CAL OUT OF TOL
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

CHECK

Remarks:

CALIBRATION OUT
OF TOLERANCE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 12:37

Test	AC	Time
01 Diagnostic	OK	12:38
02 Room Air	0.000	12:38
03 Subject Test 1	0.000	12:38
04 Room Air	0.000	12:44
05 Std. Gas	0.081	12:45
06 Room Air	0.000	12:46
07 *Subject Test	RFI*	12:47
08 Room Air	0.000	12:47

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female .Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI DETECTED

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 12:48

Test	AC	Time
01 Diagnostic	OK	12:49
02 Room Air	0.000	12:50
03 *Subject Test	0.000*	12:53
04 Room Air	0.000	12:55
05 Std. Gas	0.080	12:56
06 Room Air	0.000	12:57
07 *Subject Test	0.000*	13:01
08 Room Air	0.000	13:01

No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = DEFICIENT SAMP
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
CHARLES EDER

Remarks: DEFICIENT
SAMPLES

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 13:03

Test	AC	Time
01 Diagnostic	OK	13:04
02 Room Air	IPS*	13:04
03 Room Air	0.000	13:05

*Improper Sample

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = IMPROPER SAMPLE
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

*IMPROPER
SAMPLE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 13:06

Test	AC	Time
01 Diagnostic	OK	13:07
02 Room Air	0.000	13:07
03 Subject Test 1	0.103*	13:08
04 Room Air	0.000	13:10
05 Std. Gas	0.081	13:11
06 Room Air	0.000	13:13
07 Subject Test 2	0.146*	13:14
08 Room Air	0.000	13:15

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = DIFF TOO GREAT
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *DIFFERENCE TOO GREAT*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 13:17


Test	AC	Time
01 Diagnostic	OK	13:18
02 Room Air	0.000	13:18
03 Subject Test 1	0.146	13:19
04 Room Air	0.000	13:21
05 Std. Gas	0.081	13:22
06 Room Air	0.000	13:24
07 Subject Test 2	0.145	13:25
08 Room Air	0.000	13:26
09 Reported AC	0.145	13:25

Difference OK
No RFI Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = CMS TEST
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 6
Expiration Date = 12/05/2016
County = 08 Oper No. = 666666

20 minute waiting period ascertained? Y
Clean Mouthpiece used and disposed? Y
Clean Mouthpiece used and disposed? Y

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *CMS TEST*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006501
Location = TOXL 8164.13.00 06/09
03/29/2016 13:28

Test	AC	Time
01 Room Air	0.000	13:29
02 Std. Gas	0.080	13:29
03 Room Air	0.000	13:30
04 Std. Gas	0.081	13:30
05 Room Air	0.000	13:31
06 Std. Gas	0.081	13:31
07 Room Air	0.000	13:32
08 Std. Gas	0.081	13:32
09 Room Air	0.000	13:33
10 Std. Gas	0.081	13:33
11 Room Air	0.000	13:34
12 Std. Gas	0.080	13:34
13 Room Air	0.000	13:35
14 Std. Gas	0.081	13:35
15 Room Air	0.000	13:35
16 Std. Gas	0.081	13:36
17 Room Air	0.000	13:36
18 Std. Gas	0.081	13:37
19 Room Air	0.000	13:37
20 Std. Gas	0.080	13:38
21 Room Air	0.000	13:38
22 Std. Gas	0.081	13:39
23 Room Air	0.000	13:39
24 Std. Gas	0.081	13:39
25 Room Air	0.000	13:40
26 Std. Gas	0.081	13:40
27 Room Air	0.000	13:41
28 Std. Gas	0.081	13:41
29 Room Air	0.000	13:42
30 Std. Gas	0.081	13:42
31 Room Air	0.000	13:43
32 Std. Gas	0.081	13:43
33 Room Air	0.000	13:44
34 Std. Gas	0.081	13:44
35 Room Air	0.000	13:45
36 Std. Gas	0.082	13:45
37 Room Air	0.000	13:46
38 Std. Gas	0.082	13:46
39 Room Air	0.000	13:47
40 Std. Gas	0.082	13:47
41 Room Air	0.000	13:47
42 Cal Check Stats:		
43 Average	0.0810	
44 Std Dev	0.0006	
45 Rel Std Dev	0.70%	

STABILITY TEST
Charles E. Edr
pg 1 of 2

Operator Signature
CHARLES EDER

Remarks:

Lot: 29614080A2

CYL: 006

Exp: 12.5.2016

Form 106-I8000

STABILITY TEST

Charles Eder

pg 24/2