



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Hinton, Rick W</u>		
Location: <u>Cass County Jail</u>		
Intoxilyzer @ Serial Number: <u>80-004946</u>		
Lot Number: <u>1540080A1</u>	Cylinder Number: <u>10</u>	Expiration Date: <u>7-1-12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.077</u> %		
Test 2: <u>0.077</u> %		
Test 3: <u>0.077</u> %		
(ATTACH TEST RECORD)		

Rick W Hinton
Chemical Test Operator Signature

3-30-11
Date

Deb Kashur
Reviewed By

14 April
Date

80-004946
P 1 of 2
P 73 of 74


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = CASS 8164.13.00 06/09
03/30/2011 05:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	05:46
02 Std. Gas	0.077	05:47
03 Room Air	0.000	05:47
04 Std. Gas	0.077	05:48
05 Room Air	0.000	05:48
06 Std. Gas	0.077	05:48
07 Room Air	0.000	05:49

Lot No = 15410080A1
Cyl No = 10
Exp Date = 07/01/2012
County = 09 Oper No. = 029920



Operator Signature
RICK W. HINTON

Remarks:

Form 106-I8000

80-004946
P. 74474
RM