


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = HARV 8164.13.00 06/09
04/01/2010 23:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:49
02 Std. Gas	0.076	23:49
03 Room Air	0.000	23:50
04 Std. Gas	0.077	23:50
05 Room Air	0.000	23:51
06 Std. Gas	0.077	23:51
07 Room Air	0.000	23:51

Lot No = 709352
Cyl No = 45
Exp Date = 12/23/2011
County = 52 Oper No. = 060901



Operator Signature
TYLER A. WOLFE

Remarks: **ACA**

Form 106-I8000

80-004935
p. 41/41 KM