



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David M. Waller</u>		
Location: <u>Dickinson Det</u>		
Intoxilyzer® Serial Number: <u>80-004936</u>		
Gas Lot Number: <u>154009001</u>	Gas Cylinder Number: <u>20</u>	Expiration Date: <u>7-1-12</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0 <u>078</u> %		
Test 2: 0 <u>079</u> %		
Test 3: 0 <u>078</u> %		
(ATTACH TEST RECORD)		

David M. Waller
Chemical Test Operator Signature

4-1-11
Date

Deb Kashner
Reviewed By

14 Apr 11
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004936
Location = DKPD 8164.13.00 06/09
04/01/2011 09:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:11
02 Std. Gas	0.078	09:11
03 Room Air	0.000	09:12
04 Std. Gas	0.079	09:12
05 Room Air	0.000	09:13
06 Std. Gas	0.078	09:13
07 Room Air	0.000	09:14

Lot No = 15410080A1
Cyl No = 20
Exp Date = 07/01/2012
County = 45 Oper No. = 068332



Operator Signature
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-004936
P. 73 of 73
JAM